

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 06/27/2019

Run Number: LB103655

BalanceID: WC SC-3

OvenID: WC Oven-2

FilterID: 14466061

ThermometerID: WET-OVEN 2

TEMP1 IN: 103 06/26/2019 16:00 TEMP OUT: 104 06/27/2019 07:00

TEMP2 IN: 104 06/27/2019 07:45 TEMP OUT: 104 06/27/2019 09:00

TEMP3 IN: 105 06/27/2019 10:05 TEMP OUT: 105 06/27/2019 11:45

TEMP3 IN: 105 06/27/2019 12:30 TEMP OUT: 105 06/27/2019 14:05

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB103655BL	LB103655BL	1.4015	1.4015	100	1.4015	1.4015	1.4015	0	0
2	LB103655BS	LB103655BS	1.3987	1.3987	100	1.449	1.449	1.449	0.0503	503
3	K3547-01	062519-JETT-E-D-B	1.3873	1.3873	400	1.4018	1.4018	1.4018	0.0145	36.3
4	K3547-01DUP	062519-JETT-E-D-BDUP	1.3966	1.3966	400	1.411	1.411	1.411	0.0144	36
5	K3547-02	062519-JETT-E-D-M	1.3895	1.3895	500	1.3991	1.3991	1.3991	0.0096	19.2
6	K3547-03	062519-JETT-E-D-S	1.3758	1.3758	500	1.3788	1.3788	1.3788	0.003	6
7	K3547-04	062519-JETT-E-U-B	1.3813	1.3813	400	1.3903	1.3903	1.3903	0.009	22.5
8	K3547-05	062519-JETT-E-U-M	1.3846	1.3846	400	1.3999	1.3999	1.3999	0.0153	38.3
9	K3547-06	062519-JETT-E-U-S	1.3872	1.3872	400	1.3918	1.3918	1.3918	0.0046	11.5
10	K3547-07	062519-TIPC-E-D-B	1.3811	1.3811	400	1.3862	1.3862	1.3862	0.0051	12.8
11	K3547-08	062519-TIPC-E-D-S	1.3922	1.3922	400	1.3982	1.3982	1.3982	0.006	15
12	K3547-09	062519-TIPC-E-U-B	1.3846	1.3846	400	1.3928	1.3928	1.3928	0.0082	20.5
13	K3547-10	062519-TIPC-E-U-S	1.3848	1.3848	400	1.3912	1.3912	1.3912	0.0064	16
14	K3548-01	062419-TIPC-E-D-M	1.3886	1.3886	400	1.4002	1.4002	1.4002	0.0116	29
15	K3548-02	062419-TIPC-E-U-M	1.3835	1.3835	400	1.3906	1.3906	1.3906	0.0071	17.8



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 06/27/2019

Run Number: LB103655

BalanceID: WC SC-3

OvenID: WC Oven-2

FilterID: 14466061

ThermometerID: WET-OVEN 2

TEMP1 IN: 103 06/26/2019 16:00 TEMP OUT: 104 06/27/2019 07:00
TEMP2 IN: 104 06/27/2019 07:45 TEMP OUT: 104 06/27/2019 09:00
TEMP3 IN: 105 06/27/2019 10:05 TEMP OUT: 105 06/27/2019 11:45
TEMP3 IN: 105 06/27/2019 12:30 TEMP OUT: 105 06/27/2019 14:05

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

2B103655

WorkList Name : TSS062719

WorkList ID : 127941

Department : Wet-Chemistry

Date : 06-27-2019 07:02:32

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/02/2019	Water	K3547-01	TSS	Cool 4 deg C	HDR103	J51	062519-JETT-E-D-B	06/25/2019	SM2540 D
07/02/2019	Water	K3547-02	TSS	Cool 4 deg C	HDR103	J51	062519-JETT-E-D-M	06/25/2019	SM2540 D
07/02/2019	Water	K3547-03	TSS	Cool 4 deg C	HDR103	J51	062519-JETT-E-D-S	06/25/2019	SM2540 D
07/02/2019	Water	K3547-04	TSS	Cool 4 deg C	HDR103	J51	062519-JETT-E-U-B	06/25/2019	SM2540 D
07/02/2019	Water	K3547-05	TSS	Cool 4 deg C	HDR103	J51	062519-JETT-E-U-M	06/25/2019	SM2540 D
07/02/2019	Water	K3547-06	TSS	Cool 4 deg C	HDR103	J51	062519-JETT-E-U-S	06/25/2019	SM2540 D
07/02/2019	Water	K3547-07	TSS	Cool 4 deg C	HDR103	J51	062519-TIPC-E-D-B	06/25/2019	SM2540 D
07/02/2019	Water	K3547-08	TSS	Cool 4 deg C	HDR103	J51	062519-TIPC-E-D-S	06/25/2019	SM2540 D
07/02/2019	Water	K3547-09	TSS	Cool 4 deg C	HDR103	J51	062519-TIPC-E-U-B	06/25/2019	SM2540 D
07/02/2019	Water	K3547-10	TSS	Cool 4 deg C	HDR103	J51	062519-TIPC-E-U-S	06/25/2019	SM2540 D
07/01/2019	Water	K3548-01	TSS	Cool 4 deg C	HDR103	J51	062419-TIPC-E-D-M	06/24/2019	SM2540 D
07/01/2019	Water	K3548-02	TSS	Cool 4 deg C	HDR103	J51	062419-TIPC-E-U-M	06/24/2019	SM2540 D

Date/Time 06/27/19 09:30
 Received by: 
 Relinquished by: 

Date/Time 06/27/19 11:00
 Received by: 
 Relinquished by: 