



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB103663

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP75841
Turbidity Calibration std, 0NTU	WP75835
Turbidity Calibration std, 20NTU	WP75837
Turbidity Calibration std, 40NTU	WP75836
Turbidity Calibration std, 1NTU	WP75840
Turbidity Calibration std, 10NTU	WP75838
Turbidity Calibration std, 5NTU	WP75839
10 NTU Standard 500 ml	W2297

Intercept: 0.0888

Slope: 0.9629

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.299	06/27/2019	11:48
2	CAL2	1	1	0.976	06/27/2019	11:50
3	CAL3	5	1	4.890	06/27/2019	11:57
4	CAL4	10	1	9.850	06/27/2019	12:05
5	CAL5	20	1	18.900	06/27/2019	12:17
6	CAL6	40	1	38.800	06/27/2019	12:26



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB103663

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.360	9.628	06/27/2019	12:35
2	ICB		1	0.280	0.199	06/27/2019	12:46
3	CCV1	10	1	9.520	9.795	06/27/2019	12:52
4	CCB1		1	0.301	0.220	06/27/2019	12:59
5	LB103663BL		1	0.300	0.219	06/27/2019	13:05
6	K3566-01		1	37.500	38.853	06/27/2019	13:16
7	K3567-01		1	18.400	19.017	06/27/2019	13:24
8	K3567-01DUP		1	18.400	19.017	06/27/2019	13:29
9	CCV2	10	1	9.480	9.753	06/27/2019	13:34
10	CCB2		1	0.296	0.215	06/27/2019	13:39

WORKLIST(Hardcopy Internal Chain)

03 103663

WorkList Name : TUR062719 WorkList ID : 127988 Department : Wet-Chemistry Date : 06-27-2019 11:14:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/28/2019	Water	K3566-01	Turbidity	Cool 4 deg C	LIRO01	J53	UDS-4-29-62619	06/26/2019	SM2130 B
06/28/2019	Water	K3567-01	Turbidity	Cool 4 deg C	LIRO01	J53	UDS-4-40-62619	06/26/2019	SM2130 B

Date/Time 06-27-19 11:30
 Received by: AG
 Relinquished by: AG

Date/Time 06-27-19 13:30
 Received by: AG
 Relinquished by: AG