


LB103669

7470A

INSTRUMENT ID : CU1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.064	0.064	130	0.000	130				
0.2	0.200	0.209	0.009	718	0.0 %	718				
2.5	2.500	2.480	-0.020	9923	0.0 %	9923				
5.0	5.000	4.958	-0.042	19965	0.0 %	19965				
7.5	7.500	7.368	-0.132	29729	0.0 %	29729				
10.0	10.000	10.121	0.121	40886	0.0 %	40886				

LB103669 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 27 Jun 2019 14:46:36

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO-0	130	-	0.0000	7470A	PPB	27 Jun 2019 14:53:13	S	Std
0.2 - 1	SO-2	718	-	0.2000	7470A	PPB	27 Jun 2019 14:55:22	S	Std
2.5 - 1	SS-2	9923	-	2.5000	7470A	PPB	27 Jun 2019 14:57:31	S	Std
5.0 - 1	SS-4	19965	-	5.0000	7470A	PPB	27 Jun 2019 14:59:40	S	Std
7.5 - 1	SS-5	29729	-	7.5000	7470A	PPB	27 Jun 2019 15:01:48	S	Std
10.0 - 1	SS-10	40886	-	10.0000	7470A	PPB	27 Jun 2019 15:13:33	S	Std
ICV14 - 1	ICV14	15018	3.7376	-	7470A	PPB	27 Jun 2019 15:16:29	U	SMPL
ICB14 - 1	ICB14	20	0.0366	-	7470A	PPB	27 Jun 2019 15:18:35	U	SMPL
CCV53 - 1	CCV53	19945	4.9534	-	7470A	PPB	27 Jun 2019 15:20:42	U	SMPL
CCB53 - 1	CCB53	22	0.0371	-	7470A	PPB	27 Jun 2019 15:22:48	U	SMPL
CRA - 1	CRA	670	0.1970	-	7470A	PPB	27 Jun 2019 15:24:54	U	SMPL
HighStd - 1	HighStd	41768	10.3386	-	7470A	PPB	27 Jun 2019 15:27:00	U	SMPL
ChkStd - 1	ChkStd	27572	6.8355	-	7470A	PPB	27 Jun 2019 15:29:07	U	SMPL
PB120972BL - 1	PBW	2	0.0322	-	7470A	PPB	27 Jun 2019 15:34:03	U	SMPL
PB120972BS - 1	LCSW	17158	4.2657	-	7470A	PPB	27 Jun 2019 15:36:09	U	SMPL
K3526-11 - 1	SB-01-COMP	1249	0.3399	-	7470A	PPB	27 Jun 2019 15:38:17	U	SMPL
K3526-12 - 1	SB-02-COMP	-69	0.0147	-	7470A	PPB	27 Jun 2019 15:40:25	U	SMPL
K3526-13 - 1	SB-03-COMP	-53	0.0186	-	7470A	PPB	27 Jun 2019 15:42:33	U	SMPL
K3526-14 - 1	SB-04-COMP	-84	0.0110	-	7470A	PPB	27 Jun 2019 15:44:41	U	SMPL
K3526-15 - 1	SB-05-COMP	488	0.1521	-	7470A	PPB	27 Jun 2019 15:46:48	U	SMPL
CCV54 - 1	CCV54	19609	4.8705	-	7470A	PPB	27 Jun 2019 15:48:55	U	SMPL
CCB54 - 1	CCB54	50	0.0440	-	7470A	PPB	27 Jun 2019 15:51:01	U	SMPL
K3528-11 - 1	SB-06-COMP	1212	0.3308	-	7470A	PPB	27 Jun 2019 15:53:08	U	SMPL
K3528-13 - 1	SB-08-COMP	310	0.1082	-	7470A	PPB	27 Jun 2019 15:57:21	U	SMPL
K3528-14 - 1	SB-09-COMP	-64	0.0159	-	7470A	PPB	27 Jun 2019 15:59:28	U	SMPL
K3528-15 - 1	SB-10-COMP	-30	0.0243	-	7470A	PPB	27 Jun 2019 16:01:35	U	SMPL
K3529-01 - 1	OIL-WATER-SEPARATOR-SEDIMENT-TANK-3	344	0.1166	-	7470A	PPB	27 Jun 2019 16:03:43	U	SMPL
K3532-06 - 1	TRIANGLE-AREA-SOIL-A	262	0.0963	-	7470A	PPB	27 Jun 2019 16:05:51	U	SMPL
K3532-08 - 1	TRIANGLE-AREA-SOIL-B	194	0.0796	-	7470A	PPB	27 Jun 2019 16:07:58	U	SMPL
K3532-10 - 1	TRIANGLE-AREA-SOIL-C	107	0.0581	-	7470A	PPB	27 Jun 2019 16:10:07	U	SMPL
K3532-10DUP - 1	TRIANGLE-AREA-SOIL-CDUP	109	0.0586	-	7470A	PPB	27 Jun 2019 16:12:14	U	SMPL
CCV55 - 1	CCV55	19711	4.8957	-	7470A	PPB	27 Jun 2019 16:14:20	U	SMPL
CCB55 - 1	CCB55	39	0.0413	-	7470A	PPB	27 Jun 2019 16:16:27	U	SMPL
K3532-10MS - 1	TRIANGLE-AREA-SOIL-CMS	16360	4.0688	-	7470A	PPB	27 Jun 2019 16:18:34	U	SMPL
K3532-10MSD - 1	TRIANGLE-AREA-SOIL-CMSD	16203	4.0300	-	7470A	PPB	27 Jun 2019 16:20:40	U	SMPL
PB120973BL - 1	PBW	-108	0.0050	-	7470A	PPB	27 Jun 2019 16:22:47	U	SMPL
PB120973BS - 1	LCSW	17010	4.2292	-	7470A	PPB	27 Jun 2019 16:24:53	U	SMPL
K3542-01 - 1	MW-6R-062519	218	0.0855	-	7470A	PPB	27 Jun 2019 16:27:01	U	SMPL
K3542-01DUP - 1	MW-6R-062519DUP	247	0.0926	-	7470A	PPB	27 Jun 2019 16:29:08	U	SMPL
K3542-01MS - 1	MW-6R-062519MS	16240	4.0392	-	7470A	PPB	27 Jun 2019 16:31:16	U	SMPL
K3542-01MSD - 1	MW-6R-062519MSD	16359	4.0685	-	7470A	PPB	27 Jun 2019 16:33:24	U	SMPL
K3542-02 - 1	MW-DUP-062519	4	0.0327	-	7470A	PPB	27 Jun 2019 16:35:32	U	SMPL
K3542-03 - 1	SW-1-062519	-95	0.0083	-	7470A	PPB	27 Jun 2019 16:37:39	U	SMPL
CCV56 - 1	CCV56	18721	4.6514	-	7470A	PPB	27 Jun 2019 16:39:45	U	SMPL
CCB56 - 1	CCB56	35	0.0403	-	7470A	PPB	27 Jun 2019 16:41:52	U	SMPL
K3542-04 - 1	MW-1-062519	22	0.0371	-	7470A	PPB	27 Jun 2019 16:43:59	U	SMPL
K3542-05 - 1	MW-10-062519	-45	0.0206	-	7470A	PPB	27 Jun 2019 16:46:05	U	SMPL
PB120972TB - 1	PB120916TB	5	0.0329	-	7470A	PPB	27 Jun 2019 16:48:12	U	SMPL
K3532-10LX5 - 1		-7	0.0300	-	7470A	PPB	27 Jun 2019 16:50:18	U	SMPL
K3532-10A - 1		16035	3.9886	-	7470A	PPB	27 Jun 2019 16:54:54	U	SMPL
K3542-01LX5 - 1		-3	0.0310	-	7470A	PPB	27 Jun 2019 16:57:01	U	SMPL
K3542-01A - 1		16365	4.0700	-	7470A	PPB	27 Jun 2019 16:59:10	U	SMPL
K3528-12 - 1	SB-07-COMP	930	0.2612	-	7470A	PPB	27 Jun 2019 17:01:25	U	SMPL
CCV57 - 1	CCV57	19513	4.8468	-	7470A	PPB	27 Jun 2019 17:03:36	U	SMPL
CCB57 - 1	CCB57	-84	0.0110	-	7470A	PPB	27 Jun 2019 17:05:43	U	SMPL