



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB103674

Slope : 98.4

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/27/2019	15:00
2	CAL2	1	Water	NA	NA	20.1	7.01	06/27/2019	15:01
3	CAL3	1	Water	NA	NA	20.1	10.03	06/27/2019	15:02
4	ICV	1	Water	NA	NA	20.2	7.00	06/27/2019	15:04
5	CCV1	1	Water	NA	NA	20.1	2.02	06/27/2019	15:05
6	K3564-01	1	Water	NA	NA	27.9	6.48	06/27/2019	15:06
7	K3564-03	1	Water	NA	NA	29.3	7.56	06/27/2019	15:08
8	K3564-04	1	Water	NA	NA	29.3	5.90	06/27/2019	15:10
9	K3564-05	1	Water	NA	NA	29.2	6.61	06/27/2019	15:11
10	K3564-06	1	Water	NA	NA	28.7	6.52	06/27/2019	15:12
11	K3564-09	1	Water	NA	NA	28.8	6.49	06/27/2019	15:14
12	K3577-07	1	Water	NA	NA	27.4	7.62	06/27/2019	15:15
13	K3577-08	1	Water	NA	NA	27.2	6.95	06/27/2019	15:16
14	K3577-08DUP	1	Water	NA	NA	27.1	6.96	06/27/2019	15:18
15	CCV2	1	Water	NA	NA	20.2	12.02	06/27/2019	15:20

WORKLIST(Hardcopy Internal Chain)

LB 103674

WorkList Name : PH K3555 WorkList ID : 128000 Department : Wet-Chemistry Date : 06-27-2019 14:49:34

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/26/2019	Water	K3555-25	pH	Cool 4 deg C	PSEG01	K53	GWMH-13	06/26/2019	9040C
06/27/2019	Water	K3564-01	pH	Cool 4 deg C	PSEG01	K52	COMP-1	06/27/2019	9040C
06/27/2019	Water	K3564-03	pH	Cool 4 deg C	PSEG01	K52	13930	06/27/2019	9040C
06/27/2019	Water	K3564-04	pH	Cool 4 deg C	PSEG01	K52	13940	06/27/2019	9040C
06/27/2019	Water	K3564-05	pH	Cool 4 deg C	PSEG01	K52	13958	06/27/2019	9040C
06/27/2019	Water	K3564-06	pH	Cool 4 deg C	PSEG01	K52	13959	06/27/2019	9040C
06/27/2019	Water	K3564-09	pH	Cool 4 deg C	PSEG01	K52	13971	06/27/2019	9040C
06/27/2019	Water	K3577-07	pH	Cool 4 deg C	PSEG01	K52	30401	06/27/2019	9040C
06/27/2019	Water	K3577-08	pH	Cool 4 deg C	PSEG01	K52	30402	06/27/2019	9040C

Date/Time 06.27.19 14:50
 Received by: JA
 Relinquished by: PC

Date/Time 06.27.19 18:20
 Received by: PC
 Relinquished by: Jo