



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB103675

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/27/2019	15:00
2	CAL2	1	Water	NA	NA	20.1	7.01	06/27/2019	15:01
3	CAL3	1	Water	NA	NA	20.1	10.03	06/27/2019	15:02
4	ICV	1	Water	NA	NA	20.2	7.00	06/27/2019	15:04
5	CCV1	1	Water	NA	NA	20.1	2.02	06/27/2019	16:45
6	K3564-07	1	Solid	20.02	20	23.3	5.90	06/27/2019	16:46
7	K3564-07DUP	1	Solid	20.01	20	23.4	5.91	06/27/2019	16:47
8	K3577-01	1	Solid	20.03	20	22.9	6.90	06/27/2019	16:48
9	K3577-03	1	Solid	20.04	20	22.6	7.03	06/27/2019	16:50
10	CCV2	1	Water	NA	NA	20.1	12.02	06/27/2019	17:00

203103675

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K3577 WorkList ID : 128002 Department : Wet-Chemistry Date : 06-27-2019 16:36:50

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/27/2019	Solid	K3564-07	pH	Cool 4 deg C	PSEG01	K52	21800	06/27/2019	9041A
06/27/2019	Solid	K3577-01	pH	Cool 4 deg C	PSEG01	K52	COMP-1	06/27/2019	9041A
06/27/2019	Solid	K3577-03	pH	Cool 4 deg C	PSEG01	K52	COMP-2	06/27/2019	9041A

Date/Time 06-27-19 16:40
Received by: Jn
Relinquished by: PL

Date/Time 06-27-19 18:00
Received by: PL
Relinquished by: Jn