



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB103689

Slope : 98.4

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/27/2019	15:00
2	CAL2	1	Water	NA	NA	20.1	7.01	06/27/2019	15:01
3	CAL3	1	Water	NA	NA	20.1	10.03	06/27/2019	15:02
4	ICV	1	Water	NA	NA	20.2	6.99	06/27/2019	15:03
5	CCV1	1	Water	NA	NA	20.1	2.01	06/27/2019	15:05
6	K3555-26	1	Water	NA	NA	20.7	6.72	06/27/2019	15:06
7	K3555-26DUP	1	Water	NA	NA	20.8	6.73	06/27/2019	15:07
8	CCV2	1	Water	NA	NA	20.2	12.01	06/27/2019	15:10

LD 103689

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K3555W WorkList ID : 127999 Department : Date : 06-27-2019 14:49:06

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/26/2019	Water	K3555-26	Corrosivity	Cool 4 deg C	PSEG01	K53	GWMH-13	06/26/2019	9040C

Date/Time 06.27.19 14:50
Received by: JD
Relinquished by: PC

Date/Time 06.27.19 18:20
Received by: PC
Relinquished by: JD