

LB103693


7470A

INSTRUMENT ID : CV1

Linear

 μ Abs.:

3911E



A= 0.0000e+000

B= 2.5576e-004

C= -2.6478e-002

Rho= 0.9999779

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.006	-0.006	79	0.000	79				
0.2	0.200	0.197	-0.003	875	0.0 %	875				
2.5	2.500	2.478	-0.022	9791	0.0 %	9791				
5.0	5.000	5.051	0.051	19851	0.0 %	19851				
7.5	7.500	7.503	0.003	29440	0.0 %	29440				
10.0	10.000	9.978	-0.022	39116	0.0 %	39116				

LB103693 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 28 Jun 2019 13:32:58					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	79	-	0.0000	7470A	PPB	28 Jun 2019 13:44:46	S	Std
0.2 - 1	50-2	875	-	0.2000	7470A	PPB	28 Jun 2019 13:46:55	S	Std
2.5 - 1	52-5	9791	-	2.5000	7470A	PPB	28 Jun 2019 13:49:04	S	Std
5.0 - 1	55-0	19851	-	5.0000	7470A	PPB	28 Jun 2019 13:51:12	S	Std
7.5 - 1	57-5	29440	-	7.5000	7470A	PPB	28 Jun 2019 13:53:22	S	Std
10.0 - 1	510-0	39116	-	10.0000	7470A	PPB	28 Jun 2019 14:01:40	S	Std
ICV16 - 1	ICV16	16152	4.1045	-	7470A	PPB	28 Jun 2019 14:04:09	U	SMPL
ICB16 - 1	ICB16	1	-0.0262	-	7470A	PPB	28 Jun 2019 14:06:16	U	SMPL
CCV62 - 1	CCV62	19831	5.0454	-	7470A	PPB	28 Jun 2019 14:08:22	U	SMPL
CCB62 - 1	CCB62	2	-0.0260	-	7470A	PPB	28 Jun 2019 14:10:29	U	SMPL
CRA - 1	CRA	840	0.1884	-	7470A	PPB	28 Jun 2019 14:12:36	U	SMPL
HighStd - 1	HighStd	39246	10.0110	-	7470A	PPB	28 Jun 2019 14:14:42	U	SMPL
ChkStd - 1	ChkStd	27422	6.9869	-	7470A	PPB	28 Jun 2019 14:16:48	U	SMPL
PB121014BL - 1	PBW	48	-0.0142	-	7470A	PPB	28 Jun 2019 14:18:56	U	SMPL
PB121014BS - 1	LCSW	17955	4.5656	-	7470A	PPB	28 Jun 2019 14:21:03	U	SMPL
K3157-12 - 1	CL-02-062619-A	49	-0.0139	-	7470A	PPB	28 Jun 2019 14:23:10	U	SMPL
K3157-12DUP - 1	CL-02-062619-ADUP	75	-0.0073	-	7470A	PPB	28 Jun 2019 14:25:18	U	SMPL
K3157-12MS - 1	CL-02-062619-AMS	16089	4.0884	-	7470A	PPB	28 Jun 2019 14:27:28	U	SMPL
K3157-12MSD - 1	CL-02-062619-AMSD	16914	4.2994	-	7470A	PPB	28 Jun 2019 14:29:38	U	SMPL
K3157-14 - 1	CL-02-062619-B	45	-0.0150	-	7470A	PPB	28 Jun 2019 14:31:46	U	SMPL
CCV63 - 1	CCV63	20030	5.0963	-	7470A	PPB	28 Jun 2019 14:33:55	U	SMPL
CCB63 - 1	CCB63	2	-0.0260	-	7470A	PPB	28 Jun 2019 14:36:02	U	SMPL
K3157-16 - 1	CL-02-062619-C	85	-0.0047	-	7470A	PPB	28 Jun 2019 14:38:09	U	SMPL
K3157-18 - 1	CL-02-062619-D	81	-0.0058	-	7470A	PPB	28 Jun 2019 14:40:16	U	SMPL
K3157-20 - 1	CL-02-062619-E	86	-0.0045	-	7470A	PPB	28 Jun 2019 14:42:22	U	SMPL
K3159-10 - 1	OR-03-062619-A	83	-0.0053	-	7470A	PPB	28 Jun 2019 14:44:29	U	SMPL
K3159-12 - 1	OR-03-062619-B	-20	-0.0316	-	7470A	PPB	28 Jun 2019 14:46:36	U	SMPL
K3507-02 - 1	NB-07-062619-A	72	-0.0081	-	7470A	PPB	28 Jun 2019 14:48:43	U	SMPL
K3507-04 - 1	NB-07-062619-B	60	-0.0111	-	7470A	PPB	28 Jun 2019 14:50:51	U	SMPL
K3509-02 - 1	SA-06-062619-A	83	-0.0053	-	7470A	PPB	28 Jun 2019 14:52:59	U	SMPL
K3509-04 - 1	SA-06-062619-B	53	-0.0129	-	7470A	PPB	28 Jun 2019 14:55:07	U	SMPL
K3541-05 - 1	UNKNOWN-3B	64	-0.0101	-	7470A	PPB	28 Jun 2019 14:57:14	U	SMPL
CCV64 - 1	CCV64	19501	4.9610	-	7470A	PPB	28 Jun 2019 14:59:21	U	SMPL
CCB64 - 1	CCB64	-33	-0.0349	-	7470A	PPB	28 Jun 2019 15:01:27	U	SMPL
K3544-01 - 1	DRUM-COMP	59	-0.0114	-	7470A	PPB	28 Jun 2019 15:03:34	U	SMPL
PB121015BL - 1	PBW	118	0.0037	-	7470A	PPB	28 Jun 2019 15:05:41	U	SMPL
PB121015BS - 1	LCSW	18053	4.5907	-	7470A	PPB	28 Jun 2019 15:07:47	U	SMPL
K3541-02 - 1	UNKNOWN-2	20	-0.0214	-	7470A	PPB	28 Jun 2019 15:09:54	U	SMPL
K3541-03 - 1	UNKNOWN-3A	65	-0.0099	-	7470A	PPB	28 Jun 2019 15:12:01	U	SMPL
K3541-03DUP - 1	UNKNOWN-3ADUP	85	-0.0047	-	7470A	PPB	28 Jun 2019 15:14:09	U	SMPL
K3541-03MS - 1	UNKNOWN-3AMS	16368	4.1598	-	7470A	PPB	28 Jun 2019 15:16:16	U	SMPL
K3541-03MSD - 1	UNKNOWN-3AMSD	16384	4.1639	-	7470A	PPB	28 Jun 2019 15:18:26	U	SMPL
K3541-04 - 1	UNKNOWN-1	309	0.0526	-	7470A	PPB	28 Jun 2019 15:20:34	U	SMPL
PB121023BL - 1	PBW	88	-0.0040	-	7470A	PPB	28 Jun 2019 15:22:41	U	SMPL
CCV65 - 1	CCV65	19483	4.9564	-	7470A	PPB	28 Jun 2019 15:24:48	U	SMPL
CCB65 - 1	CCB65	13	-0.0232	-	7470A	PPB	28 Jun 2019 15:26:54	U	SMPL
PB121023BS - 1	LCSW	17771	4.5186	-	7470A	PPB	28 Jun 2019 15:29:02	U	SMPL
K3158-08 - 1	OK-02-062719	-87	-0.0487	-	7470A	PPB	28 Jun 2019 15:31:08	U	SMPL
K3158-08DUP - 1	OK-02-062719DUP	-39	-0.0365	-	7470A	PPB	28 Jun 2019 15:33:15	U	SMPL
K3158-08MS - 1	OK-02-062719MS	15648	3.9756	-	7470A	PPB	28 Jun 2019 15:35:22	U	SMPL
K3158-08MSD - 1	OK-02-062719MSD	15495	3.9365	-	7470A	PPB	28 Jun 2019 15:37:29	U	SMPL
K3163-14 - 1	SU-01-062719-A	-74	-0.0454	-	7470A	PPB	28 Jun 2019 15:39:37	U	SMPL
K3163-16 - 1	SU-01-062719-B	-84	-0.0480	-	7470A	PPB	28 Jun 2019 15:41:45	U	SMPL
K3554-04 - 1	SAMPLE-11	-54	-0.0403	-	7470A	PPB	28 Jun 2019 15:43:53	U	SMPL
K3554-08 - 1	SAMPLE-15	-37	-0.0359	-	7470A	PPB	28 Jun 2019 15:46:01	U	SMPL
K3554-12 - 1	SAMPLE-16	-30	-0.0342	-	7470A	PPB	28 Jun 2019 15:48:08	U	SMPL
CCV66 - 1	CCV66	19827	5.0444	-	7470A	PPB	28 Jun 2019 15:50:14	U	SMPL
CCB66 - 1	CCB66	30	-0.0188	-	7470A	PPB	28 Jun 2019 15:52:21	U	SMPL
K3555-04 - 1	SAMPLE-17	-77	-0.0462	-	7470A	PPB	28 Jun 2019 15:54:28	U	SMPL
K3555-08 - 1	SAMPLE-18	-66	-0.0434	-	7470A	PPB	28 Jun 2019 15:56:35	U	SMPL
K3555-12 - 1	SAMPLE-9	-61	-0.0421	-	7470A	PPB	28 Jun 2019 15:58:42	U	SMPL
K3555-16 - 1	SAMPLE-12	-45	-0.0380	-	7470A	PPB	28 Jun 2019 16:00:49	U	SMPL
K3555-20 - 1	SAMPLE-13	-67	-0.0436	-	7470A	PPB	28 Jun 2019 16:02:57	U	SMPL
K3555-24 - 1	SAMPLE-14	-66	-0.0434	-	7470A	PPB	28 Jun 2019 16:05:04	U	SMPL
K3563-02 - 1	SOIL-COMP-1	-62	-0.0423	-	7470A	PPB	28 Jun 2019 16:07:12	U	SMPL
K3563-04 - 1	SOIL-COMP-2	-66	-0.0434	-	7470A	PPB	28 Jun 2019 16:09:20	U	SMPL
PB121014TB - 1	PB120954TB	-64	-0.0428	-	7470A	PPB	28 Jun 2019 16:11:28	U	SMPL
PB121015TB - 1	PB120956TB	-64	-0.0428	-	7470A	PPB	28 Jun 2019 16:13:35	U	SMPL
CCV67 - 1	CCV67	19815	5.0414	-	7470A	PPB	28 Jun 2019 16:15:42	U	SMPL
CCB67 - 1	CCB67	-5	-0.0278	-	7470A	PPB	28 Jun 2019 16:17:49	U	SMPL
PB121023TB - 1	PB120993TB	74	-0.0076	-	7470A	PPB	28 Jun 2019 16:19:56	U	SMPL
K3157-12LX5 - 1		-33	-0.0349	-	7470A	PPB	28 Jun 2019 16:22:04	U	SMPL
K3157-12A - 1		16764	4.2610	-	7470A	PPB	28 Jun 2019 16:24:11	U	SMPL
K3541-03LX5 - 1		-59	-0.0416	-	7470A	PPB	28 Jun 2019 16:26:18	U	SMPL
K3541-03A - 1		16682	4.2401	-	7470A	PPB	28 Jun 2019 16:28:26	U	SMPL
K3158-08LX5 - 1		-81	-0.0472	-	7470A	PPB	28 Jun 2019 16:30:33	U	SMPL
K3158-08A - 1		15616	3.9674	-	7470A	PPB	28 Jun 2019 16:32:41	U	SMPL
CCV68 - 1	CCV68	19948	5.0754	-	7470A	PPB	28 Jun 2019 16:34:49	U	SMPL
CCB68 - 1	CCB68	-72	-0.0449	-	7470A	PPB	28 Jun 2019 16:36:57	U	SMPL