



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB103701

Slope : 97.8

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/28/2019	14:30
2	CAL2	1	Water	NA	NA	20.1	7.00	06/28/2019	14:31
3	CAL3	1	Water	NA	NA	20.1	10.03	06/28/2019	14:33
4	ICV	1	Water	NA	NA	20.2	7.01	06/28/2019	14:35
5	CCV1	1	Water	NA	NA	20.2	2.01	06/28/2019	14:36
6	K3585-12	1	Water	NA	NA	25.3	8.97	06/28/2019	14:37
7	K3589-01	1	Water	NA	NA	23.6	6.85	06/28/2019	14:40
8	K3589-01DUP	1	Water	NA	NA	23.7	6.86	06/28/2019	14:41
9	CCV2	1	Water	NA	NA	21.2	12.01	06/28/2019	14:43

WORKLIST(Hardcopy Internal Chain)

VP 103701

WorkList Name : PH K3589 WorkList ID : 128052 Department : Wet-Chemistry Date : 06-28-2019 14:22:03

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/28/2019	Water - ✓	K3585-12	pH	Cool 4 deg C	PSEG01	K63	RM-WATER-COMP	06/28/2019	9040C
06/28/2019	Water - ✓	K3589-01	pH	Cool 4 deg C	PSEG01	K62	FM1180	06/28/2019	9040C

Date/Time 06.28.19 14:25
 Received by: JD
 Relinquished by: LC

Date/Time 06.28.19 15:30
 Received by: LC
 Relinquished by: JD