



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB103710

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	06/28/2019	12:05
2	K3541-02		1	8	100.00	>212.0	>212.0	06/28/2019	12:35
3	K3541-02DUP		1	8	100.00	>212.0	>212.0	06/28/2019	12:55
4	K3541-03		1	9	100.00	>212.0	>212.0	06/28/2019	13:20
5	K3541-04		1	9	100.00	>212.0	>212.0	06/28/2019	13:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP62719 WorkList ID : 127968 Department : Wet-Chemistry Date : 06-27-2019 09:09:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/08/2019	Water	K3541-02 ✓	Flash Point	Cool 4 deg C	WARR01	J51	UNKNOWN-2	06/25/2019	1010A
07/08/2019	Water	K3541-03 ✓	Flash Point	Cool 4 deg C	WARR01	J51	UNKNOWN-3A	06/25/2019	1010A
07/08/2019	Water	K3541-04 ✓	Flash Point	Cool 4 deg C	WARR01	J51	UNKNOWN-1	06/25/2019	1010A

Date/Time 06/28/19 11:30
 Received by: KS
 Relinquished by: CP

Date/Time 06/28/19 13:55
 Received by: CP
 Relinquished by: KS