



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB103724

Slope : 98.1

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/01/2019	12:20
2	CAL2	1	Water	NA	NA	20.2	6.99	07/01/2019	12:21
3	CAL3	1	Water	NA	NA	20.2	10.03	07/01/2019	12:22
4	ICV	1	Water	NA	NA	20.1	7.00	07/01/2019	12:24
5	CCV1	1	Water	NA	NA	20.1	2.02	07/01/2019	12:25
6	K3619-01	1	Water	NA	NA	23.5	5.68	07/01/2019	12:26
7	K3619-01DUP	1	Water	NA	NA	23.6	5.69	07/01/2019	12:28
8	CCV2	1	Water	NA	NA	20.1	12.01	07/01/2019	12:30

WORKLIST(Hardcopy Internal Chain)

2B103724

WorkList Name : PH K3619 **WorkList ID :** 128119 **Department :** Wet-Chemistry **Date :** 07-01-2019 12:13:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/28/2019	Water	K3619-01-17	pH	Cool 4 deg C	STER01	L11	POOL-WATER	06/28/2019	SM 4500-H B

Date/Time 07-01-19 12:15
 Received by: Jg
 Relinquished by: CS

Date/Time 07-01-19 13:00
 Received by: CS
 Relinquished by: Jg