



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB103732

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	07/01/2019	14:40
2	CAL2	1	Water	NA	NA	20.1	7.01	07/01/2019	14:41
3	CAL3	1	Water	NA	NA	20.1	10.02	07/01/2019	14:42
4	ICV	1	Water	NA	NA	20.1	7.00	07/01/2019	14:43
5	CCV1	1	Water	NA	NA	20.1	2.02	07/01/2019	14:45
6	K3628-09	1	Water	NA	NA	21.7	7.09	07/01/2019	14:46
7	K3628-09DUP	1	Water	NA	NA	21.8	7.10	07/01/2019	14:47
8	CCV2	1	Water	NA	NA	20.2	12.02	07/01/2019	14:49

LB 103732

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K3628 WorkList ID : 128126 Department : Wet-Chemistry Date : 07-01-2019 14:30:57

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/01/2019	Water	K3628-09-A	pH	Cool 4 deg C	PSEG01	K61	A5551	07/01/2019	9040C

Date/Time 07-01-14 14.135
Received by: Jn
Relinquished by: [Signature]

Date/Time 07-01-14 15.00
Received by: [Signature]
Relinquished by: Jn