



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB103741

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/01/2019	17:10
2	CAL2	1	Water	NA	NA	20.2	7.01	07/01/2019	17:10
3	CAL3	1	Water	NA	NA	20.2	10.03	07/01/2019	17:12
4	ICV	1	Water	NA	NA	20.2	6.99	07/01/2019	17:15
5	CCV1	1	Water	NA	NA	20.2	2.02	07/01/2019	17:16
6	K3606-05	1	Solid	20.03	20	22.1	7.84	07/01/2019	17:18
7	K3606-05DUP	1	Solid	20.04	20	22.2	7.85	07/01/2019	17:20
8	K3628-12	1	Solid	20.01	20	20.9	7.34	07/01/2019	17:21
9	K3632-01	1	Solid	20.03	20	20.6	6.86	07/01/2019	17:22
10	K3634-03	1	Solid	20.04	20	22.0	8.15	07/01/2019	17:25
11	CCV2	1	Water	NA	NA	20.2	12.02	07/01/2019	17:26

WORKLIST(Hardcopy Internal Chain)

28 10374)

WorkList Name : CORROSIONITY K3632 WorkList ID : 128125 Department : Wet-Chemistry Date : 07-01-2019 14:30:27

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/28/2019	Solid	K3606-05	Corrosivity	Cool 4 deg C	L1R001		HJDC-COMP-1	06/28/2019	9045D
07/01/2019	Solid	K3628-12	Corrosivity	Cool 4 deg C	PSEG01	K61	DRUM-COMP	07/01/2019	9045D
06/26/2019	Solid	K3632-01	Corrosivity	Cool 4 deg C	CLEA11	L11	OIL-WATER-SEPERATOR 4-(	06/26/2019	9045D
07/01/2019	Solid	K3634-03	Corrosivity	Cool 4 deg C	L1R001	L12	HJOC-COMP-2	07/01/2019	9045D

Date/Time 07-01-19 16.00
 Received by: JH CD
 Relinquished by: JH CD

Date/Time 07-01-19 17.40
 Received by: CD
 Relinquished by: JH