



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB103749

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.6	07/02/2019	13:50
2	K3564-11		1	7	56.00	132.8	132.8	07/02/2019	14:20
3	K3564-11DUP		1	7	56.00	132.8	132.8	07/02/2019	14:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB103749

WorkList Name : FP7119 WorkList ID : 128097 Department : Wet-Chemistry Date : 07-01-2019 08:55:06

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/04/2019	Solid	K3564-11 B	Flash Point	Cool 4 deg C	PSEG01		21975	06/27/2019	1010A

Date/Time 07/02/19 9:30
 Received by: KS
 Relinquished by: CP

Date/Time 07/02/19 14:50
 Received by: CP
 Relinquished by: KS