



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB103750

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	07/02/2019	09:35
2	K3555-26		1	8	100.00	>212.0	>212.0	07/02/2019	10:10
3	K3555-26DUP		1	8	100.00	>212.0	>212.0	07/02/2019	10:40
4	K3564-01		1	9	100.00	>212.0	>212.0	07/02/2019	11:15
5	K3564-09		1	7	100.00	>212.0	>212.0	07/02/2019	11:45
6	K3577-05		1	9	100.00	>212.0	>212.0	07/02/2019	12:20
7	K3585-13		1	8	100.00	>212.0	>212.0	07/02/2019	12:45
8	K3628-09		1	7	100.00	>212.0	>212.0	07/02/2019	13:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 103750

WorkList Name : FP7119-1 WorkList ID : 128098 Department : Wet-Chemistry Date : 07-01-2019 08:56:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/09/2019	Water	K3555-26 C	Flash Point	Cool 4 deg C	PSEG01	K53	GWMH-13	06/26/2019	1010A
07/04/2019	Water	K3564-01 A	Flash Point	Cool 4 deg C	PSEG01	K52	COMP-1	06/27/2019	1010A
07/04/2019	Water	K3564-09 C	Flash Point	Cool 4 deg C	PSEG01	K52	13971	06/27/2019	1010A
07/04/2019	Water	K3577-05 A	Flash Point	Cool 4 deg C	PSEG01	K52	COMP-3	06/27/2019	1010A
07/05/2019	Water	K3585-13 B	Flash Point	Cool 4 deg C	PSEG01	K63	RM-WATER-COMP	06/28/2019	1010A
07/08/2019	Water	K3628-09 C	Flash Point	Cool 4 deg C	PSEG01	K61	A5551	07/01/2019	1010A

Date/Time 07/02/19 09:05
 Received by: KS
 Relinquished by: CP

Date/Time 07/02/19 13:25
 Received by: CP
 Relinquished by: KS