

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus, TotalSUPERVISOR REVIEW BY: apatelRun Number: LB103759BALANCE ID: WC SC-4

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP75751
calibration std. phosphate 1 ppm	WP75936
calibration std. phosphate 0.5 ppm	WP75935
calibration std. phosphate 0.3 ppm	WP75934
calibration std. 0 ppm	WP75931
phosphate ICV std.	WP75940
phosphate CCV std.	WP75939
calibration std. phosphate 0.1 ppm	WP75933
calibration std. phosphate 0.05 ppm	WP75932
Combined reagent	WP75938
11 N sulfuric acid	WP71598
Phenolphthalein indicator	WP71599
Sodium hydroxide, 1N	WP71600
PHOSPHATE LOD STD, 0.01PPM	WP75942
Phosphphate LOQ std, 0.05PPM	WP75945

Intercept: -0.0008Slope: 0.6492Regression: 0.999845

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	07/02/2019	11:42
2	CAL2	0.05	1	50	50	0.033	07/02/2019	11:42
3	CAL3	0.10	1	50	50	0.065	07/02/2019	11:43
4	CAL4	0.30	1	50	50	0.186	07/02/2019	11:44
5	CAL5	0.50	1	50	50	0.329	07/02/2019	11:45
6	CAL6	1.00	1	50	50	0.648	07/02/2019	11:46



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus, Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB103759

BALANCE ID: WC SC-4

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.325	0.502	07/02/2019	11:46
2	ICB		1	50	50	0.000	0.001	07/02/2019	11:47
3	CCV1	0.50	1	50	50	0.326	0.503	07/02/2019	12:08
4	CCB1		1	50	50	0.000	0.001	07/02/2019	12:08
5	PB121101BL		1	1.01	50	0.000	0.001	07/02/2019	12:09
6	PB121101BS	0.50	1	1.00	50	0.316	0.488	07/02/2019	12:09
7	K3591-01		1	1.01	50	0.012	0.020	07/02/2019	12:10
8	K3591-02		1	1.00	50	0.037	0.058	07/02/2019	12:10
9	CCV2	0.50	1	50	50	0.323	0.499	07/02/2019	12:11
10	CCB2		1	50	50	0.000	0.001	07/02/2019	12:12