


LB103769

7470A

INSTRUMENT ID : CU1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.001	0.001	21	0.000	21				
0.2	0.200	0.191	-0.009	775	0.0 %	775				
2.5	2.500	2.498	-0.002	9913	0.0 %	9913				
5.0	5.000	4.999	-0.001	19818	0.0 %	19818				
7.5	7.500	7.537	0.037	29870	0.0 %	29870				
10.0	10.000	9.973	-0.027	39522	0.0 %	39522				

LB103769 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 02 Jul 2019 13:20:09

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	21	-	0.0000	7470A	PPB	02 Jul 2019 13:21:38	S	Std
0.2 - 1	50.2	775	-	0.2000	7470A	PPB	02 Jul 2019 13:23:47	S	Std
2.5 - 1	52.5	9913	-	2.5000	7470A	PPB	02 Jul 2019 13:25:56	S	Std
5.0 - 1	55.0	19818	-	5.0000	7470A	PPB	02 Jul 2019 13:28:04	S	Std
7.5 - 1	57.5	29870	-	7.5000	7470A	PPB	02 Jul 2019 13:30:12	S	Std
10.0 - 1	510.0	39522	-	10.0000	7470A	PPB	02 Jul 2019 13:32:20	S	Std
ICV18 - 1	ICV18	16225	4.0919	-	7470A	PPB	02 Jul 2019 13:36:34	U	SMPL
ICB18 - 1	ICB18	-22	-0.0097	-	7470A	PPB	02 Jul 2019 13:38:40	U	SMPL
CCV73 - 1	CCV73	19979	5.0396	-	7470A	PPB	02 Jul 2019 13:40:48	U	SMPL
CCB73 - 1	CCB73	-63	-0.0201	-	7470A	PPB	02 Jul 2019 13:42:54	U	SMPL
CRA - 1	CRA	770	0.1902	-	7470A	PPB	02 Jul 2019 13:45:00	U	SMPL
HighStd - 1	HighStd	39594	9.9915	-	7470A	PPB	02 Jul 2019 13:47:06	U	SMPL
ChkStd - 1	ChkStd	27795	7.0128	-	7470A	PPB	02 Jul 2019 13:49:13	U	SMPL
PB121102BL - 1	PBW	-62	-0.0198	-	7470A	PPB	02 Jul 2019 13:51:20	U	SMPL
PB121102BS - 1	LCSW	16379	4.1308	-	7470A	PPB	02 Jul 2019 13:53:28	U	SMPL
K3158-10 - 1	OK-01-062819	-50	-0.0168	-	7470A	PPB	02 Jul 2019 13:55:35	U	SMPL
K3158-10DUP - 1	OK-01-062819DUP	-9	-0.0065	-	7470A	PPB	02 Jul 2019 13:57:43	U	SMPL
K3158-10MS - 1	OK-01-062819MS	16189	4.0828	-	7470A	PPB	02 Jul 2019 13:59:52	U	SMPL
K3158-10MSD - 1	OK-01-062819MSD	16684	4.2078	-	7470A	PPB	02 Jul 2019 14:01:59	U	SMPL
K3160-08 - 1	EO-01-062819-A	-30	-0.0118	-	7470A	PPB	02 Jul 2019 14:04:06	U	SMPL
CCV74 - 1	CCV74	20027	5.0517	-	7470A	PPB	02 Jul 2019 14:06:13	U	SMPL
CCB74 - 1	CCB74	-45	-0.0155	-	7470A	PPB	02 Jul 2019 14:08:21	U	SMPL
K3160-10 - 1	EO-01-062819-B	10	-0.0017	-	7470A	PPB	02 Jul 2019 14:10:29	U	SMPL
K3160-12 - 1	EO-01-062819-C	4	-0.0032	-	7470A	PPB	02 Jul 2019 14:12:35	U	SMPL
K3580-11 - 1	SB-11-COMP	-1	-0.0044	-	7470A	PPB	02 Jul 2019 14:14:42	U	SMPL
K3580-12 - 1	SB-12-COMP	-9	-0.0065	-	7470A	PPB	02 Jul 2019 14:16:49	U	SMPL
K3580-13 - 1	SB-13-COMP	0	-0.0042	-	7470A	PPB	02 Jul 2019 14:18:56	U	SMPL
K3580-14 - 1	SB-14-COMP	-25	-0.0105	-	7470A	PPB	02 Jul 2019 14:21:05	U	SMPL
K3580-15 - 1	SB-15-COMP	-23	-0.0100	-	7470A	PPB	02 Jul 2019 14:23:13	U	SMPL
K3592-02 - 1	RT-5124	-22	-0.0097	-	7470A	PPB	02 Jul 2019 14:25:21	U	SMPL
K3592-04 - 1	RBR-200052	-16	-0.0082	-	7470A	PPB	02 Jul 2019 14:27:30	U	SMPL
K3601-02 - 1	RT-4651	-18	-0.0087	-	7470A	PPB	02 Jul 2019 14:29:37	U	SMPL
CCV75 - 1	CCV75	19681	4.9644	-	7470A	PPB	02 Jul 2019 14:31:43	U	SMPL
CCB75 - 1	CCB75	-35	-0.0130	-	7470A	PPB	02 Jul 2019 14:33:50	U	SMPL
K3601-04 - 1	RBR-200027	3	-0.0034	-	7470A	PPB	02 Jul 2019 14:35:57	U	SMPL
PB121103BL - 1	PBW	-1	-0.0044	-	7470A	PPB	02 Jul 2019 14:38:03	U	SMPL
PB121103BS - 1	LCSW	16097	4.0596	-	7470A	PPB	02 Jul 2019 14:40:10	U	SMPL
K3555-26 - 1	GWMH-13	-25	-0.0105	-	7470A	PPB	02 Jul 2019 14:42:17	U	SMPL
K3585-13 - 1	RM-WATER-COMP	6	-0.0027	-	7470A	PPB	02 Jul 2019 14:44:25	U	SMPL
K3585-13DUP - 1	RM-WATER-COMPDUP	4	-0.0032	-	7470A	PPB	02 Jul 2019 14:46:32	U	SMPL
K3585-13MS - 1	RM-WATER-COMPMS	15570	3.9265	-	7470A	PPB	02 Jul 2019 14:48:40	U	SMPL
K3585-13MSD - 1	RM-WATER-COMPMSD	16173	4.0788	-	7470A	PPB	02 Jul 2019 14:50:49	U	SMPL
PB121104BL - 1	PBW	-26	-0.0108	-	7470A	PPB	02 Jul 2019 14:52:57	U	SMPL
PB121104BS - 1	LCSW	16134	4.0689	-	7470A	PPB	02 Jul 2019 14:55:04	U	SMPL
CCV76 - 1	CCV76	20155	5.0841	-	7470A	PPB	02 Jul 2019 14:57:11	U	SMPL
CCB76 - 1	CCB76	-31	-0.0120	-	7470A	PPB	02 Jul 2019 14:59:18	U	SMPL
K3555-25 - 1	GWMH-13	4875	1.2265	-	7470A	PPB	02 Jul 2019 15:01:24	U	SMPL
K3555-25DUP - 1	GWMH-13DUP	4228	1.0632	-	7470A	PPB	02 Jul 2019 15:03:31	U	SMPL
K3555-25MS - 1	GWMH-13MS	21173	5.3411	-	7470A	PPB	02 Jul 2019 15:05:38	U	SMPL
K3555-25MSD - 1	GWMH-13MSD	22609	5.7036	-	7470A	PPB	02 Jul 2019 15:07:44	U	SMPL
K3565-15 - 1	FL-ER-2019-1	-117	-0.0337	-	7470A	PPB	02 Jul 2019 15:09:52	U	SMPL
K3585-12 - 1	RM-WATER-COMP	-3	-0.0049	-	7470A	PPB	02 Jul 2019 15:12:00	U	SMPL
K3628-09 - 1	A5551	-62	-0.0198	-	7470A	PPB	02 Jul 2019 15:14:09	U	SMPL
PB121105BL - 1	PBW	-29	-0.0115	-	7470A	PPB	02 Jul 2019 15:16:16	U	SMPL
PB121105BS - 1	LCSW	15655	3.9480	-	7470A	PPB	02 Jul 2019 15:18:25	U	SMPL
K3606-05 - 1	HJDC-COMP-1	-87	-0.0262	-	7470A	PPB	02 Jul 2019 15:20:32	U	SMPL
CCV77 - 1	CCV77	19993	5.0432	-	7470A	PPB	02 Jul 2019 15:22:39	U	SMPL
CCB77 - 1	CCB77	-50	-0.0168	-	7470A	PPB	02 Jul 2019 15:24:46	U	SMPL
K3606-05DUP - 1	HJDC-COMP-1DUP	-52	-0.0173	-	7470A	PPB	02 Jul 2019 15:26:55	U	SMPL
K3606-05MS - 1	HJDC-COMP-1MS	15979	4.0298	-	7470A	PPB	02 Jul 2019 15:29:04	U	SMPL
K3606-05MSD - 1	HJDC-COMP-1MSD	16101	4.0606	-	7470A	PPB	02 Jul 2019 15:31:11	U	SMPL
K3628-02 - 1	VNJ-217-368	-99	-0.0292	-	7470A	PPB	02 Jul 2019 15:33:19	U	SMPL
K3628-04 - 1	VNJ-226-VNJ-238	-78	-0.0239	-	7470A	PPB	02 Jul 2019 15:35:27	U	SMPL
K3628-06 - 1	VNJ-231-VNJ-236-345	-69	-0.0216	-	7470A	PPB	02 Jul 2019 15:37:35	U	SMPL
K3632-01 - 1	OIL-WATER-SEPERATOR-4(OWS)-4	-52	-0.0173	-	7470A	PPB	02 Jul 2019 15:39:42	U	SMPL
K3634-03 - 1	HJOC-COMP-2	-63	-0.0201	-	7470A	PPB	02 Jul 2019 15:41:51	U	SMPL
PB121102TB - 1	PB121030TB	5	-0.0029	-	7470A	PPB	02 Jul 2019 15:44:00	U	SMPL
PB121103TB - 1	PB121032TB	13	-0.0009	-	7470A	PPB	02 Jul 2019 15:46:08	U	SMPL
CCV78 - 1	CCV78	20413	5.1492	-	7470A	PPB	02 Jul 2019 15:48:15	U	SMPL
CCB78 - 1	CCB78	-101	-0.0297	-	7470A	PPB	02 Jul 2019 15:50:23	U	SMPL
PB121105TB - 1	PB121085TB	-63	-0.0201	-	7470A	PPB	02 Jul 2019 15:52:30	U	SMPL
K3158-10LX5 - 1		-17	-0.0085	-	7470A	PPB	02 Jul 2019 15:54:37	U	SMPL
K3158-10A - 1		16223	4.0914	-	7470A	PPB	02 Jul 2019 15:56:44	U	SMPL
K3585-13LX5 - 1		-59	-0.0191	-	7470A	PPB	02 Jul 2019 15:58:51	U	SMPL
K3585-13A - 1		15825	3.9909	-	7470A	PPB	02 Jul 2019 16:00:58	U	SMPL
K3555-25LX5 - 1		989	0.2455	-	7470A	PPB	02 Jul 2019 16:03:07	U	SMPL
K3555-25A - 1		22177	5.5945	-	7470A	PPB	02 Jul 2019 16:05:14	U	SMPL
K3606-05LX5 - 1		-51	-0.0171	-	7470A	PPB	02 Jul 2019 16:18:09	U	SMPL
K3606-05A - 1		15994	4.0336	-	7470A	PPB	02 Jul 2019 16:20:18	U	SMPL
CCV79 - 1	CCV79	19995	5.0437	-	7470A	PPB	02 Jul 2019 16:22:26	U	SMPL
CCB79 - 1	CCB79	-40	-0.0143	-	7470A	PPB	02 Jul 2019 16:24:34	U	SMPL