



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB103778

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/02/2019	17:10
2	CAL2	1	Water	NA	NA	20.2	7.00	07/02/2019	17:11
3	CAL3	1	Water	NA	NA	20.2	10.03	07/02/2019	17:12
4	ICV	1	Water	NA	NA	20.2	7.01	07/02/2019	17:14
5	CCV1	1	Water	NA	NA	20.2	2.02	07/02/2019	17:15
6	K3650-01	1	Solid	20.02	20	23.0	6.34	07/02/2019	17:16
7	K3650-01DUP	1	Solid	20.03	20	23.1	6.35	07/02/2019	17:17
8	CCV2	1	Water	NA	NA	20.1	12.02	07/02/2019	17:20

WORKLIST(Hardcopy Internal Chain)

LB 103778

WorkList Name : PH K3650 WorkList ID : 128188 Department : Wet-Chemistry Date : 07-02-2019 16:40:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/02/2019	Solid	K3650-01	pH	Cool 4 deg C	TULL02	L12	FK-PS-01-051	07/02/2019	9045D

Date/Time 07-02-19 16:40
 Received by: JS
 Relinquished by: CS

Date/Time 07-02-19 18:00
 Received by: JS
 Relinquished by: JS