



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB103779

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/02/2019	17:10
2	CAL2	1	Water	NA	NA	20.2	7.00	07/02/2019	17:11
3	CAL3	1	Water	NA	NA	20.2	10.03	07/02/2019	17:12
4	ICV	1	Water	NA	NA	20.2	7.01	07/02/2019	17:14
5	CCV1	1	Water	NA	NA	20.2	2.02	07/02/2019	17:25
6	K3641-01	1	Water	NA	NA	22.7	7.94	07/02/2019	17:26
7	K3641-01DUP	1	Water	NA	NA	22.8	7.95	07/02/2019	17:27
8	CCV2	1	Water	NA	NA	20.1	12.02	07/02/2019	17:29

LB 103719

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K3641 WorkList ID : 128187 Department : Wet-Chemistry Date : 07-02-2019 16:39:23

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/02/2019	Water	K3641-01-X	pH	Cool 4 deg C	PSEG01	K13	FM-5210	07/02/2019	9040C

Date/Time 07-02-19 16:40
Received by: JP
Relinquished by: CD

Date/Time 07-02-19 18:00
Received by: CD
Relinquished by: JP