



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB103781

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/02/2019	17:10
2	CAL2	1	Water	NA	NA	20.2	7.00	07/02/2019	17:11
3	CAL3	1	Water	NA	NA	20.2	10.03	07/02/2019	17:12
4	ICV	1	Water	NA	NA	20.2	7.01	07/02/2019	17:14
5	CCV1	1	Water	NA	NA	20.2	2.01	07/02/2019	17:35
6	K3618-02	1	Solid	20.02	20	23.1	9.30	07/02/2019	17:36
7	K3618-02DUP	1	Solid	20.01	20	23.2	9.31	07/02/2019	17:37
8	K3648-01	1	Solid	20.03	20	24.3	9.64	07/02/2019	17:38
9	K3648-02	1	Solid	20.04	20	23.5	8.22	07/02/2019	17:40
10	CCV2	1	Water	NA	NA	20.1	12.02	07/02/2019	17:42

WORKLIST(Hardcopy Internal Chain)

103781

WorkList Name : CORROSIVITY K3618 WorkList ID : 128185 Department : Wet-Chemistry Date : 07-02-2019 16:38:04

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/02/2019	Solid	K3618-02	Corrosivity	Cool 4 deg C	PSEG05	K13	HD-01-070219	07/02/2019	9045D
07/02/2019	Solid	K3648-01	Corrosivity	Cool 4 deg C	WARR01	L41	C-1-CORE-HOLE-1	07/02/2019	9045D
07/02/2019	Solid	K3648-02	Corrosivity	Cool 4 deg C	WARR01	L41	C-2-CORE-HOLE-2	07/02/2019	9045D

Date/Time 07-02-19 16:40
 Received by: JD
 Relinquished by: JD

Date/Time 07-02-19 18:15
 Received by: CD
 Relinquished by: JD