


LB103T92

7471B

INSTRUMENT ID : CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.000	0.000	-74	0.000	-74				
0.2	0.200	0.196	-0.004	985	0.0 %	985				
2.5	2.500	2.513	0.013	13530	0.0 %	13530				
5.0	5.000	4.952	-0.048	26739	0.0 %	26739				
7.5	7.500	7.573	0.073	40930	0.0 %	40930				
10.0	10.000	9.965	-0.035	53878	0.0 %	53878				

LB103792 INSTRUMENT ID : CV1

Method: 7471B	Operator: Admin	Date of Analysis: 03 Jul 2019 10:32:38					Date	Type	Type
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units			
0.0 - 1	50.0	-74	-	0.00007471B	PPB		03 Jul 2019 10:34:20	S	Std
0.2 - 1	50.2	985	-	0.20007471B	PPB		03 Jul 2019 10:36:38	S	Std
2.5 - 1	52.5	13530	-	2.50007471B	PPB		03 Jul 2019 10:38:56	S	Std
5.0 - 1	55.0	26739	-	5.00007471B	PPB		03 Jul 2019 10:41:14	S	Std
7.5 - 1	57.5	40930	-	7.50007471B	PPB		03 Jul 2019 10:43:32	S	Std
10.0 - 1	510.0	53878	-	10.00007471B	PPB		03 Jul 2019 10:50:37	S	Std
ICV19 - 1	ICV19	23252	4.3085	-7471B	PPB		03 Jul 2019 10:55:02	U	SMPL
ICB19 - 1	ICB19	-124	-0.0088	-7471B	PPB		03 Jul 2019 10:57:18	U	SMPL
CCV80 - 1	CCV80	26434	4.8961	-7471B	PPB		03 Jul 2019 10:59:34	U	SMPL
CCB80 - 1	CCB80	-164	-0.0161	-7471B	PPB		03 Jul 2019 11:01:50	U	SMPL
CRA - 1	CRA	954	0.1903	-7471B	PPB		03 Jul 2019 11:04:06	U	SMPL
HighStd - 1	HighStd	53960	9.9798	-7471B	PPB		03 Jul 2019 11:06:21	U	SMPL
ChkStd - 1	ChkStd	38196	7.0684	-7471B	PPB		03 Jul 2019 11:08:37	U	SMPL
PB121141BL - 1	PBS	-162	-0.0158	-7471B	PPB		03 Jul 2019 11:10:55	U	SMPL
PB121141BS - 1	LCSS	20525	3.8048	-7471B	PPB		03 Jul 2019 11:13:14	U	SMPL
K3618-01 - 1	HD-01-070219	1153	0.2271	-7471B	PPB		03 Jul 2019 11:15:31	U	SMPL
K3637-01 - 1	RO-276	292	0.0681	-7471B	PPB		03 Jul 2019 11:17:48	U	SMPL
K3645-01 - 1	FK-0D015-03-010	1419	0.2762	-7471B	PPB		03 Jul 2019 11:20:06	U	SMPL
K3646-01 - 1	FK-0D015-03-011	1011	0.2009	-7471B	PPB		03 Jul 2019 11:22:23	U	SMPL
K3648-01 - 1	C-1-CORE-HOLE-1	1591	0.3080	-7471B	PPB		03 Jul 2019 11:24:39	U	SMPL
CCV81 - 1	CCV81	26457	4.9004	-7471B	PPB		03 Jul 2019 11:26:54	U	SMPL
CCB81 - 1	CCB81	-100	-0.0043	-7471B	PPB		03 Jul 2019 11:29:10	U	SMPL
K3648-02 - 1	C-2-CORE-HOLE-2	1286	0.2517	-7471B	PPB		03 Jul 2019 11:31:27	U	SMPL
K3648-02DUP - 1	C-2-CORE-HOLE-2DUP	1252	0.2454	-7471B	PPB		03 Jul 2019 11:33:47	U	SMPL
K3648-02MS - 1	C-2-CORE-HOLE-2MS	21430	3.9720	-7471B	PPB		03 Jul 2019 11:36:06	U	SMPL
K3648-02MSD - 1	C-2-CORE-HOLE-2MSD	21884	4.0558	-7471B	PPB		03 Jul 2019 11:38:23	U	SMPL
K3649-01 - 1	D-1-ROLL-OFF-SOCK-LILE	5319	0.9965	-7471B	PPB		03 Jul 2019 11:40:39	U	SMPL
K3650-01 - 1	FK-PS-01-051	389	0.0860	-7471B	PPB		03 Jul 2019 11:42:56	U	SMPL
K3648-02LX5 - 1		10	0.0160	-7471B	PPB		03 Jul 2019 11:45:13	U	SMPL
K3648-02A - 1		21799	4.0401	-7471B	PPB		03 Jul 2019 11:47:30	U	SMPL
CCV82 - 1	CCV82	26371	4.8845	-7471B	PPB		03 Jul 2019 11:49:47	U	SMPL
CCB82 - 1	CCB82	-100	-0.0043	-7471B	PPB		03 Jul 2019 11:52:03	U	SMPL