


LB103795

7470A

INSTRUMENT ID : CUI



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.002	0.002	-8	0.000	-8				
0.2	0.200	0.192	-0.008	730	0.0 %	730				
2.5	2.500	2.485	-0.015	9641	0.0 %	9641				
5.0	5.000	5.048	0.048	19607	0.0 %	19607				
7.5	7.500	7.473	-0.027	29033	0.0 %	29033				
10.0	10.000	10.000	-0.000	38853	0.0 %	38853				

LB103795 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 03 Jul 2019 12:15:26

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	-8	-	0.0000	7470A	PPB	03 Jul 2019 12:15:37	S	Std
0.2 - 1	50.2	730	-	0.2000	7470A	PPB	03 Jul 2019 12:17:46	S	Std
2.5 - 1	52.5	9641	-	2.5000	7470A	PPB	03 Jul 2019 12:19:54	S	Std
5.0 - 1	55.0	19607	-	5.0000	7470A	PPB	03 Jul 2019 12:22:03	S	Std
7.5 - 1	57.5	29033	-	7.5000	7470A	PPB	03 Jul 2019 12:24:11	S	Std
10.0 - 1	510.0	38853	-	10.0000	7470A	PPB	03 Jul 2019 12:35:56	S	Std
ICV20 - 1	ICV20	15846	4.0808	-	7470A	PPB	03 Jul 2019 12:41:04	U	SMPL
ICB20 - 1	ICB20	-57	-0.0104	-	7470A	PPB	03 Jul 2019 12:43:11	U	SMPL
CCV83 - 1	CCV83	19786	5.0944	-	7470A	PPB	03 Jul 2019 12:45:18	U	SMPL
CCB83 - 1	CCB83	-73	-0.0145	-	7470A	PPB	03 Jul 2019 12:47:25	U	SMPL
CRA - 1	CRA	697	0.1836	-	7470A	PPB	03 Jul 2019 12:49:32	U	SMPL
HighStd - 1	HighStd	38754	9.9741	-	7470A	PPB	03 Jul 2019 12:51:39	U	SMPL
ChkStd - 1	ChkStd	27260	7.0172	-	7470A	PPB	03 Jul 2019 12:53:46	U	SMPL
PB121149BL - 1	PBW	-93	-0.0196	-	7470A	PPB	03 Jul 2019 12:55:53	U	SMPL
PB121149BS - 1	LCSW	16153	4.1598	-	7470A	PPB	03 Jul 2019 12:58:00	U	SMPL
K3618-02 - 1	HD-01-070219	-60	-0.0112	-	7470A	PPB	03 Jul 2019 13:00:09	U	SMPL
K3618-02DUP - 1	HD-01-070219DUP	-26	-0.0024	-	7470A	PPB	03 Jul 2019 13:02:17	U	SMPL
K3618-02MSD - 1	HD-01-070219MSD	16015	4.1243	-	7470A	PPB	03 Jul 2019 13:06:33	U	SMPL
K3649-03 - 1	D-1-ROLL-OFF-SOCK-LILE	-38	-0.0055	-	7470A	PPB	03 Jul 2019 13:08:40	U	SMPL
CCV84 - 1	CCV84	19749	5.0849	-	7470A	PPB	03 Jul 2019 13:10:47	U	SMPL
CCB84 - 1	CCB84	-74	-0.0148	-	7470A	PPB	03 Jul 2019 13:12:54	U	SMPL
PB121150BL - 1	PBW	-39	-0.0058	-	7470A	PPB	03 Jul 2019 13:15:00	U	SMPL
PB121150BS - 1	LCSW	16279	4.1922	-	7470A	PPB	03 Jul 2019 13:17:07	U	SMPL
K3519-02 - 1	SP-19-LF2MW4	-63	-0.0119	-	7470A	PPB	03 Jul 2019 13:19:13	U	SMPL
K3519-03 - 1	SP-19-LF2MW4-9001	-19	-0.0006	-	7470A	PPB	03 Jul 2019 13:21:20	U	SMPL
K3519-09 - 1	SP-19-LF2MW3	-26	-0.0024	-	7470A	PPB	03 Jul 2019 13:23:28	U	SMPL
K3636-11 - 1	SP19-MMW0018B	-45	-0.0073	-	7470A	PPB	03 Jul 2019 13:25:36	U	SMPL
K3638-01 - 1	MWHR2-6-070119	-39	-0.0058	-	7470A	PPB	03 Jul 2019 13:27:44	U	SMPL
K3638-01DUP - 1	MWHR2-6-070119DUP	-33	-0.0042	-	7470A	PPB	03 Jul 2019 13:29:53	U	SMPL
K3638-02 - 1	K3638-01MS	15995	4.1191	-	7470A	PPB	03 Jul 2019 13:41:53	U	SMPL
K3638-03 - 1	K3638-01MSD	15951	4.1078	-	7470A	PPB	03 Jul 2019 13:44:00	U	SMPL
CCV85 - 1	CCV85	19831	5.1060	-	7470A	PPB	03 Jul 2019 13:46:07	U	SMPL
CCB85 - 1	CCB85	-78	-0.0158	-	7470A	PPB	03 Jul 2019 13:48:15	U	SMPL
K3638-04 - 1	DUP	-34	-0.0045	-	7470A	PPB	03 Jul 2019 13:50:21	U	SMPL
K3638-05 - 1	MWHR2-10-070119	-33	-0.0042	-	7470A	PPB	03 Jul 2019 13:52:29	U	SMPL
K3641-01 - 1	FM-5210	-12	0.0012	-	7470A	PPB	03 Jul 2019 13:54:35	U	SMPL
PB121149TB - 1	PB121126TB	-19	-0.0006	-	7470A	PPB	03 Jul 2019 13:56:42	U	SMPL
K3618-02LX5 - 1		0	0.0043	-	7470A	PPB	03 Jul 2019 13:58:50	U	SMPL
K3618-02A - 1		16023	4.1263	-	7470A	PPB	03 Jul 2019 14:06:02	U	SMPL
K3638-01LX5 - 1		-44	-0.0070	-	7470A	PPB	03 Jul 2019 14:08:09	U	SMPL
K3638-01A - 1		15836	4.0782	-	7470A	PPB	03 Jul 2019 14:10:20	U	SMPL
K3618-02MS - 1	HD-01-070219MS	16158	4.1611	-	7470A	PPB	03 Jul 2019 14:12:28	U	SMPL
CCV86 - 1	CCV86	19641	5.0571	-	7470A	PPB	03 Jul 2019 14:14:35	U	SMPL
CCB86 - 1	CCB86	-71	-0.0140	-	7470A	PPB	03 Jul 2019 14:16:43	U	SMPL