



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB103800

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	07/03/2019	09:45
2	K3641-02		1	8	100.00	>212.0	>212.0	07/03/2019	10:20
3	K3641-02DUP		1	8	100.00	>212.0	>212.0	07/03/2019	10:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB103800

WorkList Name : FP7319 **WorkList ID :** 128218 **Department :** Wet-Chemistry **Date :** 07-03-2019 09:14:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/15/2019	Water	K3641-02	C Flash Point	Cool 4 deg C	PSEG01	K13	FM-5210	07/02/2019	1010A

Date/Time 07/03/19 09:15
Received by: KS
Relinquished by: CP

Date/Time 07/03/19 10:55
Received by: CP
Relinquished by: KS