


LB103801

ISM02.4

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.4222e-004

C= 3.8450e-002

Rho= 0.9999411

Accept=Accepted

slope
yintercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.042	0.042	15	0.000	15					—
0.05	0.050										
0.20	0.200	0.221	0.021	754	0.0 %	754					11
2.50	2.500	2.457	-0.043	9986	0.0 %	9986					-2
5.00	5.000	4.934	-0.066	20211	0.0 %	20211					-1
7.50	7.500	7.513	0.013	30858	0.0 %	30858					0
10.0	10.000	10.033	0.033	41260	0.0 %	41260					0

ps 07.03.19

LB103801
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	15	-		0 ISM02.4	PPB	7/3/2019 14:35	Std
	0.2 S0.2	754	-		0.2 ISM02.4	PPB	7/3/2019 14:37	Std
	2.5 S2.5	9986	-		2.5 ISM02.4	PPB	7/3/2019 14:39	Std
	5 S5.0	20211	-		5 ISM02.4	PPB	7/3/2019 14:41	Std
	7.5 S7.5	30858	-		7.5 ISM02.4	PPB	7/3/2019 14:44	Std
	10 S10.0	41260	-		10 ISM02.4	PPB	7/3/2019 15:00	Std
ICV	ICV	16014	3.9174	-	ISM02.4	PPB	7/3/2019 15:02	SMPL
ICB	ICB	-50	0.0263	-	ISM02.4	PPB	7/3/2019 15:04	SMPL
CCV087	CCV087	20308	4.9575	-	ISM02.4	PPB	7/3/2019 15:07	SMPL
CCB087	CCB087	-52	0.0259	-	ISM02.4	PPB	7/3/2019 15:09	SMPL
PB121154BL	PBW154	3	0.0392	-	ISM02.4	PPB	7/3/2019 15:11	SMPL
K3521-07	SY-6	-25	0.0324	-	ISM02.4	PPB	7/3/2019 15:14	SMPL
K3521-08	SY-3DD	-7	0.0368	-	ISM02.4	PPB	7/3/2019 15:16	SMPL
K3521-09	SY-3D	155	0.076	-	ISM02.4	PPB	7/3/2019 15:18	SMPL
K3521-48	PK-10S	-41	0.0285	-	ISM02.4	PPB	7/3/2019 15:20	SMPL
K3521-49	PK-10D	23	0.044	-	ISM02.4	PPB	7/3/2019 15:23	SMPL
K3521-50	PK-10I	1823	0.48	-	ISM02.4	PPB	7/3/2019 15:25	SMPL
K3521-51	K3521-50D	1781	0.4698	-	ISM02.4	PPB	7/3/2019 15:27	SMPL
K3521-52	K3521-50S	6310	1.5669	-	ISM02.4	PPB	7/3/2019 15:30	SMPL
K3521-53	RW-12D	0	0.0385	-	ISM02.4	PPB	7/3/2019 15:32	SMPL
K3521-54	RW-12I	7	0.0401	-	ISM02.4	PPB	7/3/2019 15:34	SMPL
K3521-55	SY-4	106	0.0641	-	ISM02.4	PPB	7/3/2019 15:36	SMPL
K3521-56	SY-3	25	0.0445	-	ISM02.4	PPB	7/3/2019 15:39	SMPL
K3521-57	SY-2R	-20	0.0336	-	ISM02.4	PPB	7/3/2019 15:41	SMPL
K3521-58	SY-2D	2	0.0389	-	ISM02.4	PPB	7/3/2019 15:43	SMPL
PB121155BL	PBW155	5	0.0397	-	ISM02.4	PPB	7/3/2019 15:45	SMPL
K3521-01	FIELD-BLANK	-23	0.0329	-	ISM02.4	PPB	7/3/2019 15:48	SMPL
K3521-02	SY-6	8	0.0404	-	ISM02.4	PPB	7/3/2019 15:50	SMPL
K3521-03	SY-3DD	-10	0.036	-	ISM02.4	PPB	7/3/2019 15:52	SMPL
K3521-04	SY-3D	26	0.0447	-	ISM02.4	PPB	7/3/2019 15:54	SMPL
CCV088	CCV088	20096	4.9062	-	ISM02.4	PPB	7/3/2019 15:57	SMPL
CCB088	CCB088	-30	0.0312	-	ISM02.4	PPB	7/3/2019 15:59	SMPL
K3521-36	PK-10S	-23	0.0329	-	ISM02.4	PPB	7/3/2019 16:01	SMPL
K3521-37	PK-10D	553	0.1724	-	ISM02.4	PPB	7/3/2019 16:04	SMPL
K3521-38	PK-10I	1641	0.4359	-	ISM02.4	PPB	7/3/2019 16:06	SMPL
K3521-41	RW-12D	71	0.0556	-	ISM02.4	PPB	7/3/2019 16:08	SMPL
K3521-42	RW-12I	49	0.0503	-	ISM02.4	PPB	7/3/2019 16:10	SMPL
K3521-44	SY-4	92	0.0607	-	ISM02.4	PPB	7/3/2019 16:13	SMPL
K3521-45	SY-3	8	0.0404	-	ISM02.4	PPB	7/3/2019 16:15	SMPL
K3521-46	SY-2R	-17	0.0343	-	ISM02.4	PPB	7/3/2019 16:17	SMPL
K3521-47	SY-2D	-20	0.0336	-	ISM02.4	PPB	7/3/2019 16:19	SMPL
K3521-59	K3521-38D	1645	0.4369	-	ISM02.4	PPB	7/3/2019 16:22	SMPL
K3521-60	K3521-38S	6165	1.5318	-	ISM02.4	PPB	7/3/2019 16:24	SMPL
PB121156BL	PBW156	-19	0.0338	-	ISM02.4	PPB	7/3/2019 16:27	SMPL
K3642-01	MCOJS8	97	0.0619	-	ISM02.4	PPB	7/3/2019 16:29	SMPL
K3642-02	MCOJTO	95	0.0615	-	ISM02.4	PPB	7/3/2019 16:31	SMPL

LB103801
INSTRUMENT ID : CV1

K3642-03	MCOJT2	104	0.0636 -	ISM02.4	PPB	7/3/2019 16:33	SMPL
K3642-04	MCOJW1	112	0.0656 -	ISM02.4	PPB	7/3/2019 16:36	SMPL
K3642-05	MCOJW2	5	0.0397 -	ISM02.4	PPB	7/3/2019 16:38	SMPL
K3642-06	MCOJW3	178	0.0816 -	ISM02.4	PPB	7/3/2019 16:40	SMPL
K3642-07	MCOJW4	112	0.0656 -	ISM02.4	PPB	7/3/2019 16:43	SMPL
K3642-08	MCOJW5	105	0.0639 -	ISM02.4	PPB	7/3/2019 16:45	SMPL
CCV089	CCV089	19915	4.8623 -	ISM02.4	PPB	7/3/2019 16:47	SMPL
CCB089	CCB089	-37	0.0295 -	ISM02.4	PPB	7/3/2019 16:49	SMPL
K3642-09	MCOJW6	69	0.0552 -	ISM02.4	PPB	7/3/2019 16:52	SMPL
K3642-10	MCOJW8	52	0.051 -	ISM02.4	PPB	7/3/2019 16:54	SMPL
K3642-11	MCOJS9	49	0.0503 -	ISM02.4	PPB	7/3/2019 16:56	SMPL
K3642-12	MCOJT1	1186	0.3257 -	ISM02.4	PPB	7/3/2019 16:59	SMPL
K3642-13	MCOJT3	36	0.0472 -	ISM02.4	PPB	7/3/2019 17:01	SMPL
K3642-14	MCOJT4	47	0.0498 -	ISM02.4	PPB	7/3/2019 17:03	SMPL
K3642-15	MCOJT5	36	0.0472 -	ISM02.4	PPB	7/3/2019 17:05	SMPL
K3642-16	MCOJT6	55	0.0518 -	ISM02.4	PPB	7/3/2019 17:08	SMPL
K3642-17	MCOJT6D	47	0.0498 -	ISM02.4	PPB	7/3/2019 17:10	SMPL
K3642-18	MCOJT6S	4231	1.0633 -	ISM02.4	PPB	7/3/2019 17:12	SMPL
K3642-19	MCOJT7	59	0.0527 -	ISM02.4	PPB	7/3/2019 17:14	SMPL
K3642-20	MCOJT8	19	0.0431 -	ISM02.4	PPB	7/3/2019 17:17	SMPL
K3642-21	MCOJT9	54	0.0515 -	ISM02.4	PPB	7/3/2019 17:19	SMPL
K3642-22	MCOJW0	1	0.0387 -	ISM02.4	PPB	7/3/2019 17:21	SMPL
PB121156TB	LEB156	171	0.0799 -	ISM02.4	PPB	7/3/2019 17:24	SMPL
CCV090	CCV090	19699	4.81 -	ISM02.4	PPB	7/3/2019 17:26	SMPL
CCB090	CCB090	-52	0.0259 -	ISM02.4	PPB	7/3/2019 17:28	SMPL