



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB103809

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/03/2019	17:40
2	CAL2	1	Water	NA	NA	20.2	7.00	07/03/2019	17:41
3	CAL3	1	Water	NA	NA	20.2	10.03	07/03/2019	17:42
4	ICV	1	Water	NA	NA	20.2	7.01	07/03/2019	17:43
5	CCV1	1	Water	NA	NA	20.2	2.02	07/03/2019	18:15
6	K3508-03	1	Solid	20.02	20	23.3	8.44	07/03/2019	18:16
7	K3508-04	1	Solid	20.01	20	22.8	8.50	07/03/2019	18:17
8	K3508-05	1	Solid	20.04	20	23.0	8.44	07/03/2019	18:18
9	K3508-06	1	Solid	20.02	20	23.0	8.65	07/03/2019	18:20
10	K3508-07	1	Solid	20.02	20	22.9	8.29	07/03/2019	18:21
11	K3508-08	1	Solid	20.03	20	22.8	8.70	07/03/2019	18:22
12	K3508-09	1	Solid	20.04	20	22.8	8.70	07/03/2019	18:24
13	K3508-10	1	Solid	20.01	20	22.9	8.73	07/03/2019	18:25
14	K3508-11	1	Solid	20.04	20	22.7	8.68	07/03/2019	18:26
15	K3508-12	1	Solid	20.02	20	22.5	8.66	07/03/2019	18:27
16	CCV2	1	Water	NA	NA	20.1	12.02	07/03/2019	18:30
17	K3508-12DUP	1	Solid	20.03	20	22.6	8.67	07/03/2019	18:32
18	CCV3	1	Water	NA	NA	20.1	2.02	07/03/2019	18:34

1-27 103809

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K3508 WorkList ID : 128242 Department : Wet-Chemistry Date : 07-03-2019 15:29:37

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/03/2019	Solid	K3508-03	pH	Cool 4 deg C	PSEG01	K12	4804	07/03/2019	9045D
07/03/2019	Solid	K3508-04	pH	Cool 4 deg C	PSEG01	K12	4805	07/03/2019	9045D
07/03/2019	Solid	K3508-05	pH	Cool 4 deg C	PSEG01	K12	4806	07/03/2019	9045D
07/03/2019	Solid	K3508-06	pH	Cool 4 deg C	PSEG01	K12	4807	07/03/2019	9045D
07/03/2019	Solid	K3508-07	pH	Cool 4 deg C	PSEG01	K12	4808	07/03/2019	9045D
07/03/2019	Solid	K3508-08	pH	Cool 4 deg C	PSEG01	K12	4809	07/03/2019	9045D
07/03/2019	Solid	K3508-09	pH	Cool 4 deg C	PSEG01	K12	4810	07/03/2019	9045D
07/03/2019	Solid	K3508-10	pH	Cool 4 deg C	PSEG01	K12	4811	07/03/2019	9045D
07/03/2019	Solid	K3508-11	pH	Cool 4 deg C	PSEG01	K12	4812	07/03/2019	9045D
07/03/2019	Solid	K3508-12	pH	Cool 4 deg C	PSEG01	K12	4813	07/03/2019	9045D

Date/Time 07-03-19 15:30
 Received by: JA
 Relinquished by: JKL

Date/Time 07-03-19 18:00
 Received by: JA
 Relinquished by: JD