

Element, Wavelength and Order	Date of Fit	Date of Cal.	Type of Fit	Weighting	A0	A1	A2	n (Exponent)	Correlation
As 189.042 {479}	07/04/2019 0:27:31	07/03/2019 11:31:42	Linear	1/Conc	-0.000691	0.062838	0.000000	1.000000	0.999794
Tl 190.856 {477}	07/04/2019 0:27:31	07/03/2019 11:31:42	Linear	1/Conc	-0.001296	0.097127	0.000000	1.000000	0.999750
Pb 220.353 {453}	07/04/2019 0:27:31	07/03/2019 11:31:42	Linear	1/Conc	-0.000210	0.201100	0.000000	1.000000	0.999996
Se 196.090 {472}	07/04/2019 0:27:31	07/03/2019 11:31:42	Linear	1/Conc	0.000823	0.074622	0.000000	1.000000	0.999854
Sb 206.833 {463}	07/04/2019 0:27:31	07/03/2019 11:31:42	Linear	1/Conc	0.001167	0.210482	0.000000	1.000000	0.999869
Al 396.152 { 85}	07/04/2019 0:27:31	07/03/2019 11:36:06	Linear	1/Conc	0.003734	0.109031	0.000000	1.000000	0.999578
Ba 493.409 { 68}	07/04/2019 0:27:31	07/03/2019 11:31:42	Linear	1/Conc	0.003348	3.776300	0.000000	1.000000	0.999831
Be 234.861 {144}	07/04/2019 0:27:31	07/03/2019 11:31:42	Linear	1/Conc	-0.003517	5.761068	0.000000	1.000000	0.999540
Cd 214.438 {457}	07/04/2019 0:27:31	07/03/2019 11:31:42	Linear	1/Conc	0.000450	2.843595	0.000000	1.000000	0.999932
Ca 373.690 { 90}	07/04/2019 0:27:31	07/03/2019 11:36:06	Linear	1/Conc	0.000512	0.066785	0.000000	1.000000	0.999789
Cr 267.716 {126}	07/04/2019 0:27:31	07/03/2019 11:31:42	Linear	1/Conc	-0.000057	0.206310	0.000000	1.000000	0.999657
Co 228.616 {448}	07/04/2019 0:27:31	07/03/2019 11:31:42	Linear	1/Conc	-0.000273	0.860752	0.000000	1.000000	0.999996
Cu 324.754 {104}	07/04/2019 0:27:31	07/03/2019 11:31:42	Linear	1/Conc	0.002245	0.223747	0.000000	1.000000	0.999892
Fe 259.837 {130}	07/04/2019 0:27:31	07/03/2019 11:36:06	Linear	1/Conc	0.000179	0.032465	0.000000	1.000000	0.999828
Mn 257.610 {131}	07/04/2019 0:27:31	07/03/2019 11:31:42	Linear	1/Conc	0.000522	0.753270	0.000000	1.000000	0.999750
Mg 279.079 {121}	07/04/2019 0:27:31	07/03/2019 11:36:06	Linear	1/Conc	-0.000088	0.014703	0.000000	1.000000	0.999705
Ni 231.604 {446}	07/04/2019 0:27:31	07/03/2019 11:31:42	Linear	1/Conc	-0.003667	0.818938	0.000000	1.000000	0.999957
Ag 328.068 {103}	07/04/2019 0:27:31	07/03/2019 11:31:42	Linear	1/Conc	-0.000857	0.274543	0.000000	1.000000	0.999934
Na 818.326 { 41}	07/04/2019 0:27:31	07/03/2019 11:36:06	Linear	1/Conc	-0.001366	0.005340	0.000000	1.000000	0.998975
V 292.402 {115}	07/04/2019 0:27:31	07/03/2019 11:31:42	Linear	1/Conc	0.000695	0.157212	0.000000	1.000000	0.999846
Zn 206.200 {464}	07/04/2019 0:27:31	07/03/2019 11:31:42	Linear	1/Conc	3.478024	1929.7410	0.000000	1.000000	0.995888
Zn 213.856 {158}	07/04/2019 0:27:31	07/03/2019 11:31:42	Linear	1/Conc	-0.000241	0.341932	0.000000	1.000000	0.999698
K 769.896 { 44}	07/04/2019 0:27:31	07/03/2019 11:31:42	Linear	1/Conc	0.023338	0.023458	0.000000	1.000000	0.999951
P 177.495 {490}	07/04/2019 0:27:31	07/03/2019 11:31:42	Linear	1/Conc	-0.001076	0.052919	0.000000	1.000000	0.999982
B 249.678 {135}	07/04/2019 0:27:31	07/03/2019 11:31:42	Linear	1/Conc	0.002916	0.332869	0.000000	1.000000	0.997225
Mo 202.030 {467}	07/04/2019 0:27:31	07/03/2019 11:31:42	Linear	1/Conc	0.000880	1.010782	0.000000	1.000000	0.999906
S 182.034 {485}	07/04/2019 0:27:31	07/03/2019 11:31:42	Linear	1/Conc	0.000326	0.029129	0.000000	1.000000	0.999901
Si 251.611 {134}	07/04/2019 0:27:31	07/03/2019 11:31:42	Linear	1/Conc	0.008652	0.404266	0.000000	1.000000	0.999661
Sn 189.989 {478}	07/04/2019 0:27:31	07/03/2019 11:31:42	Linear	1/Conc	0.000309	0.154719	0.000000	1.000000	0.999997
Ti 336.121 {100}	07/04/2019 0:27:31	07/03/2019 11:31:42	Linear	1/Conc	0.002272	0.535960	0.000000	1.000000	0.999871
Li 670.784 { 50}	07/04/2019 0:27:31	07/03/2019 11:31:42	Linear	1/Conc	0.013101	0.898037	0.000000	1.000000	0.997915
Y 224.306 {150}*	<not fit>	<Never Calibrated>	Linear	1/Conc	0.000000	0.000000	0.000000	1.000000	0.000000
Y 224.306 {450}*	<not fit>	<Never Calibrated>	Linear	1/Conc	0.000000	0.000000	0.000000	1.000000	0.000000
Y 360.073 { 94}*	<not fit>	<Never Calibrated>	Linear	1/Conc	0.000000	0.000000	0.000000	1.000000	0.000000
Y 371.030 { 91}*	<not fit>	<Never Calibrated>	Linear	1/Conc	0.000000	0.000000	0.000000	1.000000	0.000000
In 230.606 {446}*	<not fit>	<Never Calibrated>	Linear	1/Conc	0.000000	0.000000	0.000000	1.000000	0.000000
Sr 407.771 { 83}	07/04/2019 0:27:31	07/03/2019 11:31:42	Linear	1/Conc	0.000186	8.268672	0.000000	1.000000	0.999833