



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB103828

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	07/08/2019	09:20
2	CAL2	1	Water	NA	NA	20.1	7.00	07/08/2019	09:21
3	CAL3	1	Water	NA	NA	20.1	10.03	07/08/2019	09:22
4	ICV	1	Water	NA	NA	20.1	6.99	07/08/2019	09:24
5	CCV1	1	Water	NA	NA	20.2	2.02	07/08/2019	09:25
6	K3688-01	1	Water	NA	NA	21.6	5.86	07/08/2019	09:26
7	K3688-01DUP	1	Water	NA	NA	21.7	5.87	07/08/2019	09:27
8	CCV2	1	Water	NA	NA	20.1	12.02	07/08/2019	09:30

23103828

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph k3688 WorkList ID : 128288 Department : Wet-Chemistry Date : 07-08-2019 09:10:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/24/2019	Water	K3688-01 - 1	pH	Cool 4 deg C	STER14	L11	JS-10000-SYSTEM	06/24/2019	SM 4500-H B

Date/Time 07-08-14 9:15
Received by: SA PC
Relinquished by: SA PC

Date/Time 07-08-19 10:10
Received by: PC
Relinquished by: SA