



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : AMANDEEP

Run Number: LB103853

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/09/2019	12:05
2	CAL2	1	Water	NA	NA	20.2	7.01	07/09/2019	12:06
3	CAL3	1	Water	NA	NA	20.2	10.03	07/09/2019	12:07
4	ICV	1	Water	NA	NA	20.2	6.99	07/09/2019	12:10
5	CCV1	1	Water	NA	NA	20.2	2.01	07/09/2019	12:11
6	K3712-02	1	Water	NA	NA	16.3	8.75	07/09/2019	12:13
7	K3712-02DUP	1	Water	NA	NA	16.4	8.76	07/09/2019	12:14
8	K3712-03	1	Water	NA	NA	17.0	8.61	07/09/2019	12:15
9	CCV2	1	Water	NA	NA	20.2	12.01	07/09/2019	12:16

LB103853

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K3712 WorkList ID : 128353 Department : Wet-Chemistry Date : 07-09-2019 11:53:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/08/2019	Water	K3712-02 - <u>1</u>	pH	Cool 4 deg C	TETR06	L32	TT-003-IDWGW-20190708	07/08/2019	9040C
07/08/2019	Water	K3712-03 - <u>1</u>	pH	Cool 4 deg C	TETR06	L32	TT-004-IDWGW-20190708	07/08/2019	9040C

Date/Time 07-09-19 12:00
Received by: jo
Relinquished by: cd

Date/Time 07-09-19 12:35
Received by: cd
Relinquished by: jo