


LB103858

7471B

INSTRUMENT ID : CU1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.046	0.046	750	0.000	750				
0.2	0.200	0.213	0.013	1663	0.0 %	1663				
2.5	2.500	2.476	-0.024	13984	0.0 %	13984				
5.0	5.000	4.952	-0.048	27465	0.0 %	27465				
7.5	7.500	7.438	-0.062	40999	0.0 %	40999				
10.0	10.000	10.075	0.075	55354	0.0 %	55354				

LB103858 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 09 Jul 2019 13:50:36					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	750	-	0.0000	7471B	PPB	09 Jul 2019 13:56:01	S	Std
0.2 - 1	50.2	1663	-	0.2000	7471B	PPB	09 Jul 2019 13:58:19	S	Std
2.5 - 1	52.5	13984	-	2.5000	7471B	PPB	09 Jul 2019 14:00:37	S	Std
5.0 - 1	55.0	27465	-	5.0000	7471B	PPB	09 Jul 2019 14:02:54	S	Std
7.5 - 1	57.5	40999	-	7.5000	7471B	PPB	09 Jul 2019 14:05:12	S	Std
10.0 - 1	50.8	55354	-	10.0000	7471B	PPB	09 Jul 2019 14:07:29	S	Std
ICV26 - 1	ICV26	21726	3.8982	-	7471B	PPB	09 Jul 2019 14:10:08	U	SMPL
ICB26 - 1	ICB26	708	0.0379	-	7471B	PPB	09 Jul 2019 14:12:24	U	SMPL
CCV03 - 1	CCV03	27568	4.9712	-	7471B	PPB	09 Jul 2019 14:14:40	U	SMPL
CCB03 - 1	CCB03	730	0.0419	-	7471B	PPB	09 Jul 2019 14:16:56	U	SMPL
CRA - 1	CRA	1706	0.2212	-	7471B	PPB	09 Jul 2019 14:19:13	U	SMPL
HighStd - 1	HighStd	55612	10.1220	-	7471B	PPB	09 Jul 2019 14:21:30	U	SMPL
ChkStd - 1	ChkStd	38460	6.9717	-	7471B	PPB	09 Jul 2019 14:23:48	U	SMPL
PB121242BL - 1	PBS	576	0.0136	-	7471B	PPB	09 Jul 2019 14:26:08	U	SMPL
PB121242BS - 1	LCSS	20499	3.6728	-	7471B	PPB	09 Jul 2019 14:28:52	U	SMPL
K3627-01 - 1	TR-05-070819	4407	0.7173	-	7471B	PPB	09 Jul 2019 14:31:09	U	SMPL
K3627-01DUP - 1	TR-05-070819DUP	4439	0.7231	-	7471B	PPB	09 Jul 2019 14:33:26	U	SMPL
K3627-01MS - 1	TR-05-070819MS	22715	4.0799	-	7471B	PPB	09 Jul 2019 14:35:43	U	SMPL
K3627-01MSD - 1	TR-05-070819MSD	24971	4.4942	-	7471B	PPB	09 Jul 2019 14:38:01	U	SMPL
K3687-03 - 1	STONES	514	0.0023	-	7471B	PPB	09 Jul 2019 14:40:17	U	SMPL
CCV04 - 1	CCV04	27316	4.9249	-	7471B	PPB	09 Jul 2019 14:42:33	U	SMPL
CCB04 - 1	CCB04	689	0.0344	-	7471B	PPB	09 Jul 2019 14:44:49	U	SMPL
K3693-01 - 1	RBR-200028	846	0.0632	-	7471B	PPB	09 Jul 2019 14:47:05	U	SMPL
K3694-01 - 1	GC-SOIL-PILE	4406	0.7171	-	7471B	PPB	09 Jul 2019 14:49:20	U	SMPL
K3697-01 - 1	72-12000	2523	0.3712	-	7471B	PPB	09 Jul 2019 14:51:36	U	SMPL
K3697-04 - 1	RBR-200035	1521	0.1872	-	7471B	PPB	09 Jul 2019 14:53:53	U	SMPL
K3699-01 - 1	FK-RF-01-014-ABCD	6009	1.0115	-	7471B	PPB	09 Jul 2019 14:56:10	U	SMPL
K3700-01 - 1	FK-PS-01-052	2346	0.3387	-	7471B	PPB	09 Jul 2019 14:58:27	U	SMPL
K3703-01 - 1	FK-EF-01-015-ABCD	4665	0.7647	-	7471B	PPB	09 Jul 2019 15:00:44	U	SMPL
K3627-01LX5 - 1		1251	0.1376	-	7471B	PPB	09 Jul 2019 15:03:01	U	SMPL
K3627-01A - 1		23634	4.2486	-	7471B	PPB	09 Jul 2019 15:05:19	U	SMPL
CCV05 - 1	CCV05	27259	4.9144	-	7471B	PPB	09 Jul 2019 15:07:35	U	SMPL
CCB05 - 1	CCB05	450	-0.0095	-	7471B	PPB	09 Jul 2019 15:09:52	U	SMPL