



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 7/10/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 17:10
In Date: 07/09/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 07/10/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB103866

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K3626-01	SU-02-070919-A	1	1.12	9.73	8.98	91.3
K3626-03	SU-02-070919-B	2	1.12	9.56	9.24	96.2
K3706-01	390702715	3	1.14	9.76	1.57	5.0
K3707-01	390703726.2	4	1.16	9.76	1.35	2.2
K3708-01	390703764-30989	5	1.15	9.96	3.61	27.9
K3709-01	390702024.0-SLUDGE-TRUCI	6	1.19	9.98	1.47	3.2
K3710-01	390703732	7	1.13	9.97	2.7	17.8
K3711-01	390703138.0-CHSTP	8	1.14	9.96	1.78	7.3
K3715-01	S14-01-B	9	1.17	9.66	9.24	95.1
K3715-02	S16-01-B	10	1.12	9.55	9.07	94.3
K3715-03	S16-01-S	11	1.11	9.6	8.25	84.1
K3715-04	S16-04-B	12	1.15	9.98	9.54	95.0
K3715-05	S16-04-S	13	1.13	9.57	9.03	93.6
K3715-06	DUP-19	14	1.18	9.94	9.6	96.1
K3716-01	CONCRETE-1	15	1.00	2.00	2.00	100.0
K3716-02	CONCRETE-2	16	1.00	2.00	2.00	100.0
K3716-03	CONCRETE-3	17	1.00	2.00	2.00	100.0
K3716-04	WHITE-BLOCK	18	1.00	2.00	2.00	100.0
K3716-05	BLUE-BLOCK	19	1.00	2.00	2.00	100.0
K3716-06	RECTANGULAR-BLOCK	20	1.00	2.00	2.00	100.0
K3716-07	SQUARE-BLOCK	21	1.00	2.00	2.00	100.0
K3721-11	13759	22	1.00	2.00	2.00	100.0
K3726-01	PX-1_070919	23	1.14	9.62	8.68	88.9
K3726-02	PX-2_070919	24	1.16	9.77	9.01	91.2
K3726-03	PX-3_070919	25	1.19	9.73	8.38	84.2
K3726-04	PX-4_070919	26	1.17	9.79	8.77	88.2
K3726-05	PIPE-1_070919	27	1.12	9.94	8.45	83.1
K3726-06	COMP-1_070919	28	1.15	9.96	8.67	85.4

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$