



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB103867

Slope : 98.2

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	07/09/2019	17:15
2	CAL2	1	Water	NA	NA	20.1	7.00	07/09/2019	17:16
3	CAL3	1	Water	NA	NA	20.1	10.03	07/09/2019	17:17
4	ICV	1	Water	NA	NA	20.1	6.99	07/09/2019	17:19
5	CCV1	1	Water	NA	NA	20.1	2.02	07/09/2019	17:20
6	K3718-02	1	Water	NA	NA	26.3	5.06	07/09/2019	17:21
7	K3718-02DUP	1	Water	NA	NA	26.4	5.07	07/09/2019	17:22
8	K3721-03	1	Water	NA	NA	28.2	4.38	07/09/2019	17:24
9	K3721-04	1	Water	NA	NA	27.0	6.76	07/09/2019	17:25
10	K3721-05	1	Water	NA	NA	26.8	5.73	07/09/2019	17:26
11	K3721-06	1	Water	NA	NA	26.6	5.51	07/09/2019	17:28
12	K3721-07	1	Water	NA	NA	27.0	5.43	07/09/2019	17:30
13	K3721-09	1	Water	NA	NA	26.1	5.18	07/09/2019	17:31
14	CCV2	1	Water	NA	NA	20.1	12.02	07/09/2019	17:33

13103867

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph k3721 WorkList ID : 128362 Department : Wet-Chemistry Date : 07-09-2019 16:45:33

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/09/2019	Water	K3718-02	-A pH	Cool 4 deg C	PSEG01	M22	30434	07/09/2019	9040C
07/09/2019	Water	K3721-03	-A pH	Cool 4 deg C	PSEG01	M22	13941	07/09/2019	9040C
07/09/2019	Water	K3721-04	-A pH	Cool 4 deg C	PSEG01	M22	13928	07/09/2019	9040C
07/09/2019	Water	K3721-05	-A pH	Cool 4 deg C	PSEG01	M22	13942	07/09/2019	9040C
07/09/2019	Water	K3721-06	-A pH	Cool 4 deg C	PSEG01	M22	30415	07/09/2019	9040C
07/09/2019	Water	K3721-07	-A pH	Cool 4 deg C	PSEG01	M22	30416	07/09/2019	9040C
07/09/2019	Water	K3721-09	-A pH	Cool 4 deg C	PSEG01	M22	50113	07/09/2019	9040C

Date/Time 07-09-19 16:50
 Received by: Jn
 Relinquished by: CP

Date/Time 07-09-19 18:20
 Received by: CD
 Relinquished by: JF