



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB103868

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	07/09/2019	17:15
2	CAL2	1	Water	NA	NA	20.1	7.00	07/09/2019	17:16
3	CAL3	1	Water	NA	NA	20.1	10.03	07/09/2019	17:17
4	ICV	1	Water	NA	NA	20.1	6.99	07/09/2019	17:19
5	CCV1	1	Water	NA	NA	20.2	2.02	07/09/2019	17:35
6	K3626-02	1	Solid	20.02	20	22.8	7.52	07/09/2019	17:36
7	K3626-02DUP	1	Solid	20.03	20	22.9	7.53	07/09/2019	17:38
8	K3626-04	1	Solid	20.04	20	22.6	8.99	07/09/2019	17:40
9	CCV2	1	Water	NA	NA	20.2	12.01	07/09/2019	17:41

1B 102868

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity k3626 WorkList ID : 128361 Department : Wet-Chemistry Date : 07-09-2019 16:44:49

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/09/2019	Solid	K3626-02	Corrosivity	Cool 4 deg C	PSEG05	M22	SU-02-070919-A	07/09/2019	9045D
07/09/2019	Solid	K3626-04	Corrosivity	Cool 4 deg C	PSEG05	M22	SU-02-070919-B	07/09/2019	9045D

Date/Time 07-09-19 1645
 Received by: J
 Relinquished by: GP

Date/Time 07-09-19 18.25
 Received by: GP
 Relinquished by: GP