



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB103872

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	07/10/2019	10:05
2	K3718-02		1	9	100.00	>212.0	>212.0	07/10/2019	10:35
3	K3718-02DUP		1	9	100.00	>212.0	>212.0	07/10/2019	10:55
4	K3721-01		1	7	100.00	>212.0	>212.0	07/10/2019	11:20
5	K3721-07		1	8	100.00	>212.0	>212.0	07/10/2019	11:50
6	K3721-09		1	6	100.00	>212.0	>212.0	07/10/2019	12:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 103872

WorkList Name : FP71019 WorkList ID : 128387 Department : Wet-Chemistry Date : 07-10-2019 09:24:28

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/10/2019	Water	K3718-02 <i>C</i>	Flash Point	Cool 4 deg C	PSEG01	M22	30434	07/09/2019	1010A
07/16/2019	Water	K3721-01 <i>A</i>	Flash Point	Cool 4 deg C	PSEG01	M22	COMP-1	07/09/2019	1010A
07/16/2019	Water	K3721-07 <i>C</i>	Flash Point	Cool 4 deg C	PSEG01	M22	30416	07/09/2019	1010A
07/16/2019	Water	K3721-09 <i>C</i>	Flash Point	Cool 4 deg C	PSEG01	M22	50113	07/09/2019	1010A

Date/Time 07/10/19 09:30
 Received by: *KS*
 Relinquished by: *CP*

Date/Time 07/10/19 12:25
 Received by: *CP*
 Relinquished by: *KS*