



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB103876

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP76161
Turbidity Calibration std, 0NTU	WP76162
Turbidity Calibration std, 20NTU	WP76157
Turbidity Calibration std, 40NTU	WP76156
Turbidity Calibration std, 1NTU	WP76160
Turbidity Calibration std, 10NTU	WP76158
Turbidity Calibration std, 5NTU	WP76159
10 NTU Standard 500 ml	W2297

Intercept: 0.1313

Slope: 0.9624

Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.290	07/10/2019	11:36
2	CAL2	1	1	0.992	07/10/2019	11:38
3	CAL3	5	1	4.880	07/10/2019	11:40
4	CAL4	10	1	9.870	07/10/2019	11:42
5	CAL5	20	1	19.200	07/10/2019	11:45
6	CAL6	40	1	38.700	07/10/2019	11:48



Analytical Summary Report

Analysis Method: SM2130 B
Parameter: Turbidity
Run Number: LB103876

ANALYST: AMANDEEP
SUPERVISOR REVIEW BY: apatel

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.500	9.735	07/10/2019	11:51
2	ICB		1	0.287	0.162	07/10/2019	11:55
3	CCV1	10	1	9.710	9.953	07/10/2019	11:58
4	CCB1		1	0.290	0.165	07/10/2019	12:05
5	LB103876BL		1	0.286	0.161	07/10/2019	12:08
6	K3591-10		1	0.528	0.412	07/10/2019	12:11
7	K3591-11		1	1.070	0.975	07/10/2019	12:16
8	K3736-01		1	33.300	34.465	07/10/2019	12:22
9	K3737-01		1	12.500	12.852	07/10/2019	12:25
10	K3737-01DUP		1	12.400	12.748	07/10/2019	12:28
11	CCV2	10	1	9.690	9.932	07/10/2019	12:31
12	CCB2		1	0.290	0.165	07/10/2019	12:34

WORKLIST(Hardcopy Internal Chain)

LB103870

WorkList Name : tur071019 WorkList ID : 128398 Department : Wet-Chemistry Date : 07-10-2019 10:47:41

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/30/2019	Water	K3591-10	Turbidity	Cool 4 deg C	CHEM02	QA Office	LOD-MDL-WATER-04-QT3-20	06/28/2019	SM2130 B
06/30/2019	Water	K3591-11	Turbidity	Cool 4 deg C	CHEM02	QA Office	LOQ-WATER-05-QT3-2019	06/28/2019	SM2130 B
07/11/2019	Water	K3736-01	Turbidity	Cool 4 deg C	LIRO01	M12	UDS-4-26-70919	07/09/2019	SM2130 B
07/11/2019	Water	K3737-01	Turbidity	Cool 4 deg C	LIRO01	M12	UDS-4-40-70919	07/09/2019	SM2130 B

Date/Time 07-10-19 11:10
 Received by: AK
 Relinquished by: AK

Date/Time 07-10-19 12:45
 Received by: AK
 Relinquished by: AK