



Analytical Summary Report

Analysis Method: 9041A

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB103893

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467
3.0-5.5 pH paper refill	W1934
Test Paper, PH, Short Range 6.0/8.0	W2350
Test Paper,PH Short Range 9.0/10.0	W1939
0.0-3.0 pH paper refill	W1931
10.0-12.0 pH paper refill	W1940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.0	07/10/2019	17:50
2	CAL2	1	Water	NA	NA	20.1	7.0	07/10/2019	17:51
3	CAL3	1	Water	NA	NA	20.1	10.0	07/10/2019	17:52
4	ICV	1	Water	NA	NA	20.1	7.0	07/10/2019	18:00
5	CCV1	1	Water	NA	NA	20.2	2.0	07/10/2019	18:01
6	K3730-03	1	Solid	20.02	20	22.9	6.0	07/10/2019	18:04
7	K3735-01	1	Solid	20.03	20	21.6	5.5	07/10/2019	18:05
8	K3735-01DUP	1	Solid	20.03	20	21.7	5.6	07/10/2019	18:10
9	CCV2	1	Water	NA	NA	20.1	12.0	07/10/2019	18:11

LB 103893

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K3735 WorkList ID : 128413 Department : Wet-Chemistry Date : 07-10-2019 17:43:16

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/10/2019	Solid	K3730-03	pH	Cool 4 deg C	PSEG01	M23	13956	07/10/2019	9041A
07/10/2019	Solid	K3735-01	pH	Cool 4 deg C	PSEG01	M31	1586	07/10/2019	9041A

Date/Time 07.10.19 17.43
Received by: JG
Relinquished by: JG

Date/Time 07.10.19 18.25
Received by: PC
Relinquished by: JG