



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB103894

Slope : 98.4

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/10/2019	18:10
2	CAL2	1	Water	NA	NA	20.2	7.01	07/10/2019	18:11
3	CAL3	1	Water	NA	NA	20.2	10.03	07/10/2019	18:12
4	ICV	1	Water	NA	NA	20.2	6.99	07/10/2019	18:15
5	CCV1	1	Water	NA	NA	20.1	2.02	07/10/2019	18:30
6	K3730-01	1	Water	NA	NA	24.2	5.62	07/10/2019	18:31
7	K3730-01DUP	1	Water	NA	NA	24.3	5.63	07/10/2019	18:33
8	CCV2	1	Water	NA	NA	20.2	12.01	07/10/2019	18:35

187 103894

WORKLIST(Hardcopy Internal Chain)

WorkList Name :	PH K3730	WorkList ID :	128414	Department :	Wet-Chemistry	Date :	07-10-2019 17:44:13		
Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/10/2019	Water	K3730-01	pH	Cool 4 deg C	PSEG01	M23	13773	07/10/2019	9040C

Date/Time 07.10.19 17:45
Received by: SP
Relinquished by: SP

Date/Time 07.10.19 18.25
Received by: PC
Relinquished by: SP