

LB103902

~~7471B~~
7470A


INSTRUMENT ID : CU1

RS 07-11-19

Linear

 μ Abs.:

44078



A= 0.0000e+000

B= 2.2562e-004

C= -5.7918e-002

Rho= 0.9997886

Accept = Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.034	-0.034	104	0.000	104				
0.2	0.200	0.168	-0.032	1002	0.0 %	1002				
2.5	2.500	2.489	-0.011	11287	0.0 %	11287				
5.0	5.000	5.120	0.120	22949	0.0 %	22949				
7.5	7.500	7.571	0.071	33814	0.0 %	33814				
10.0	10.000	9.887	-0.113	44078	0.0 %	44078				

LB103902 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 11 Jul 2019 11:23:52					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	104	-	0.0000	7471B	PPB	11 Jul 2019 11:27:22	S	Std
0.2 - 1	50.2	1002	-	0.2000	7471B	PPB	11 Jul 2019 11:29:40	S	Std
2.5 - 1	52.5	11287	-	2.5000	7471B	PPB	11 Jul 2019 11:31:58	S	Std
5.0 - 1	55.0	22949	-	5.0000	7471B	PPB	11 Jul 2019 11:36:48	S	Std
7.5 - 1	57.5	33814	-	7.5000	7471B	PPB	11 Jul 2019 11:39:05	S	Std
10.0 - 1	510.0	44078	-	10.0000	7471B	PPB	11 Jul 2019 11:41:25	S	Std
ICV29 - 1	ICV29	18105	4.0269	-	7471B	PPB	11 Jul 2019 11:44:03	U	SMPL
ICB29 - 1	ICB29	-101	-0.0807	-	7471B	PPB	11 Jul 2019 11:46:20	U	SMPL
CCV13 - 1	CCV13	23327	5.2050	-	7471B	PPB	11 Jul 2019 11:54:10	U	SMPL
CCB13 - 1	CCB13	-105	-0.0816	-	7471B	PPB	11 Jul 2019 11:56:25	U	SMPL
CRA - 1	CRA	896	0.1442	-	7471B	PPB	11 Jul 2019 11:58:43	U	SMPL
HighStd - 1	HighStd	45826	10.2812	-	7471B	PPB	11 Jul 2019 12:00:58	U	SMPL
ChkStd - 1	ChkStd	30742	6.8780	-	7471B	PPB	11 Jul 2019 12:03:15	U	SMPL
PB121307BL - 1	PBW	-47	-0.0685	-	7471B	PPB	11 Jul 2019 12:05:34	U	SMPL
PB121307BS - 1	LCSW	19076	4.2460	-	7471B	PPB	11 Jul 2019 12:07:52	U	SMPL
K3626-02 - 1	SU-02-070919-A	11	-0.0554	-	7471B	PPB	11 Jul 2019 12:10:09	U	SMPL
K3626-02DUP - 1	SU-02-070919-ADUP	132	-0.0281	-	7471B	PPB	11 Jul 2019 12:12:27	U	SMPL
K3626-02MS - 1	SU-02-070919-AMS	19569	4.3572	-	7471B	PPB	11 Jul 2019 12:14:45	U	SMPL
K3626-02MSD - 1	SU-02-070919-AMSD	20615	4.5932	-	7471B	PPB	11 Jul 2019 12:17:04	U	SMPL
K3626-04 - 1	SU-02-070919-B	-74	-0.0746	-	7471B	PPB	11 Jul 2019 12:19:22	U	SMPL
CCV14 - 1	CCV14	21582	4.8113	-	7471B	PPB	11 Jul 2019 12:21:40	U	SMPL
CCB14 - 1	CCB14	-10	-0.0602	-	7471B	PPB	11 Jul 2019 12:23:58	U	SMPL
PB121307TB - 1	PB121246TB	16	-0.0543	-	7471B	PPB	11 Jul 2019 12:26:16	U	SMPL
K3626-02LX5 - 1		64	-0.0435	-	7471B	PPB	11 Jul 2019 12:28:31	U	SMPL
K3626-02A - 1		20200	4.4995	-	7471B	PPB	11 Jul 2019 12:30:47	U	SMPL
CCV15 - 1	CCV15	20834	4.6426	-	7471B	PPB	11 Jul 2019 12:33:04	U	SMPL
CCB15 - 1	CCB15	-95	-0.0794	-	7471B	PPB	11 Jul 2019 12:35:21	U	SMPL