



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB103904

Slope : 98.7

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/11/2019	11:30
2	CAL2	1	Water	NA	NA	20.2	7.00	07/11/2019	11:31
3	CAL3	1	Water	NA	NA	20.2	10.04	07/11/2019	11:32
4	ICV	1	Water	NA	NA	20.2	6.99	07/11/2019	11:35
5	CCV1	1	Water	NA	NA	20.2	2.01	07/11/2019	11:36
6	K3758-01	1	Water	NA	NA	22.4	4.91	07/11/2019	11:38
7	K3758-01DUP	1	Water	NA	NA	22.5	4.92	07/11/2019	11:39
8	K3759-01	1	Water	NA	NA	22.4	5.22	07/11/2019	11:42
9	CCV2	1	Water	NA	NA	20.2	12.02	07/11/2019	11:44

2-13103964

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K3758 WorkList ID : 128459 Department : Wet-Chemistry Date : 07-11-2019 11:23:05

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/10/2019	Water	K3758-01 - 15	pH	Cool 4 deg C	STER14	M21	JS-10000-SYSTEM	07/10/2019	SM 4500-H B
07/05/2019	Water	K3759-01 - 15	pH	Cool 4 deg C	STER11	M11	POOLWATER	07/05/2019	SM 4500-H B

Date/Time 07-11-19 11:25
Received by: gn
Relinquished by: pc

Date/Time 07-11-19 12:00
Received by: pc
Relinquished by: gn