



## Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB103922

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

| Reagent/Standard | Lot/Log # |
|------------------|-----------|
| p-xylene (ICV)   | W2168     |

| Seq | LabID       | True Value °F | DL | Initial Sample °C | Celsius °C | Result °F | Final Result °F | Anal Date  | Anal Time |
|-----|-------------|---------------|----|-------------------|------------|-----------|-----------------|------------|-----------|
| 1   | ICV         | 81            | 1  | 7                 | 27.00      | 80.6      | 80.6            | 07/11/2019 | 10:10     |
| 2   | K3730-01    |               | 1  | 8                 | 100.00     | >212.0    | >212.0          | 07/11/2019 | 10:40     |
| 3   | K3730-01DUP |               | 1  | 8                 | 100.00     | >212.0    | >212.0          | 07/11/2019 | 11:15     |

Result = (Celsius \* 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) \* 0.06

# WORKLIST(Hardcopy Internal Chain)

LB103922

WorkList Name : FP71119      WorkList ID : 128446      Department : Wet-Chemistry      Date : 07-11-2019 09:27:42

| Due Date   | Matrix | Sample   | Test        | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method |
|------------|--------|----------|-------------|--------------|----------|------------------|-----------------|--------------|--------|
| 07/17/2019 | Water  | K3730-01 | Flash Point | Cool 4 deg C | PSEG01   | M23              | 13773           | 07/10/2019   | 1010A  |

Date/Time 07/11/19 09:20  
 Received by: KS  
 Relinquished by: CP

Date/Time 07/11/19 11:20  
 Received by: CP  
 Relinquished by: KS