



Analytical Summary Report

Analysis Method: 9041A

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB103933

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467
3.0-5.5 pH paper refill	W1934
Test Paper, PH, Short Range 6.0/8.0	W2350
Test Paper,PH Short Range 9.0/10.0	W1939
0.0-3.0 pH paper refill	W1931
10.0-12.0 pH paper refill	W1940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.0	07/12/2019	13:15
2	CAL2	1	Water	NA	NA	21.3	7.0	07/12/2019	13:16
3	CAL3	1	Water	NA	NA	21.3	10.0	07/12/2019	13:18
4	ICV	1	Water	NA	NA	21.3	7.0	07/12/2019	13:20
5	CCV1	1	Water	NA	NA	21.4	2.0	07/12/2019	13:22
6	K3784-01	1	Solid	20.02	20	22.5	9.5	07/12/2019	13:25
7	K3784-01DUP	1	Solid	20.01	20	22.6	9.7	07/12/2019	13:26
8	CCV2	1	Water	NA	NA	21.4	12.0	07/12/2019	13:28

18103933

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH S K3784 WorkList ID : 128507 Department : Wet-Chemistry Date : 07-12-2019 13:06:26

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/12/2019	Solid	K3784-01-15	pH	Cool 4 deg C	PSEG01	M42	OILY-DEBRIS	07/12/2019	9041A

Date/Time 07.12.19 13:10
Received by: JN
Relinquished by: PL

Date/Time 07.12.19 14:00
Received by: PL
Relinquished by: JN