

213/03938

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

Instrument ID : Konelab

7/12/2019 15:33

Test: Total CN

Ng/L

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	94.167	0.0	0.083	
ICB1	0.536	0.0	0.002	
CCV1	247.539	0.0	0.217	
CCB1	0.112	0.0	0.001	
PB121236BL	0.269	0.0	0.001	
PB121236BS	95.522	0.0	0.084	
MIDPB121236	247.620	0.0	0.217	
K3591-12	2.071	0.0	0.003	
CCV2	247.527	0.0	0.217	
CCB2	0.483	0.0	0.002	

N	10
Mean	93.584
SD	112.6828
CV%	120.41

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

Instrument ID : Konelab

7/12/2019 15:11


Test Total CN

Accepted 7/12/2019 11:33

Factor 1148
Bias 0.001

Coeff. of det. 0.999963

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.001	-0.6859	0.0000	
2	5.0PPBCN	0.006	4.9129	5.0000	
3	10PPBCN	0.010	10.1639	10.0000	
4	50PPBCN	0.046	51.1243	50.0000	
5	100PPBCN	0.087	98.4261	100.0000	
6	250PPBCN	0.220	251.7174	250.0000	
7	500PPBCN	0.436	499.3412	500.0000	

Aquakem v. 7.2AQ1

Results from time period:

Fri Jul 12 15:25:55 2019

Fri Jul 12 15:26:05 2019

AS

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	-0.6859	µg/l	7/12/2019 11:30:14	
5.0PPBCN	A	Total CN	P	4.9129	µg/l	7/12/2019 11:30:15	
10PPBCN	A	Total CN	P	10.1639	µg/l	7/12/2019 11:30:16	
50PPBCN	A	Total CN	P	51.1243	µg/l	7/12/2019 11:30:17	
100PPBCN	A	Total CN	P	98.4261	µg/l	7/12/2019 11:30:18	
250PPBCN	A	Total CN	P	251.7174	µg/l	7/12/2019 11:30:19	
500PPBCN	A	Total CN	P	499.3412	µg/l	7/12/2019 11:30:20	
ICV1	S	Total CN	P	94.1669	µg/l	7/12/2019 15:25:55	
ICB1	S	Total CN	P	0.5355	µg/l	7/12/2019 15:25:56	
CCV1	S	Total CN	P	247.5386	µg/l	7/12/2019 15:25:57	
CCB1	S	Total CN	P	0.1115	µg/l	7/12/2019 15:25:58	
PB121236BL	S	Total CN	P	0.2685	µg/l	7/12/2019 15:25:59	
PB121236BS	S	Total CN	P	95.5224	µg/l	7/12/2019 15:26:00	
MIDPB121236	S	Total CN	P	247.6198	µg/l	7/12/2019 15:26:01	
K3591-12	S	Total CN	P	2.0708	µg/l	7/12/2019 15:26:03	
CCV2	S	Total CN	P	247.5266	µg/l	7/12/2019 15:26:04	
CCB2	S	Total CN	P	0.4833	µg/l	7/12/2019 15:26:05	