



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB103946

Slope : 98.3

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	22.3	4.01	07/12/2019	17:20
2	CAL2	1	Water	NA	NA	22.3	7.00	07/12/2019	17:21
3	CAL3	1	Water	NA	NA	22.3	10.03	07/12/2019	17:22
4	ICV	1	Water	NA	NA	22.3	6.99	07/12/2019	17:24
5	CCV1	1	Water	NA	NA	22.3	2.02	07/12/2019	17:25
6	K3788-01	1	Water	NA	NA	27.2	7.50	07/12/2019	17:26
7	K3788-01DUP	1	Water	NA	NA	27.3	7.51	07/12/2019	17:27
8	CCV2	1	Water	NA	NA	22.3	12.02	07/12/2019	17:30

28103946

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K3788 WorkList ID : 128519 Department : Wet-Chemistry Date : 07-12-2019 16:41:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/12/2019	Water	K3788-01- 2	pH	Cool 4 deg C	TULL02	M51	2056	07/12/2019	9040C

Date/Time 07-12-19 16:50
 Received by: JA
 Relinquished by: cd

Date/Time 07-12-19 18:10
 Received by: cd
 Relinquished by: JA