



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB103947

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	22.3	4.01	07/12/2019	17:20
2	CAL2	1	Water	NA	NA	22.3	7.00	07/12/2019	17:21
3	CAL3	1	Water	NA	NA	22.3	10.03	07/12/2019	17:22
4	ICV	1	Water	NA	NA	22.3	6.99	07/12/2019	17:24
5	CCV1	1	Water	NA	NA	22.3	2.01	07/12/2019	17:35
6	K3807-01	1	Solid	20.03	20	23.9	6.76	07/12/2019	17:36
7	K3807-01DUP	1	Solid	20.04	20	23.8	6.77	07/12/2019	17:37
8	CCV2	1	Water	NA	NA	22.4	12.02	07/12/2019	17:39

✓ 103947

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K3807 WorkList ID : 128522 Department : Wet-Chemistry Date : 07-12-2019 17:08:32

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/12/2019	Solid	K3807-01	pH	Cool 4 deg C	TULL02	M31	FK-PS-01-053	07/12/2019	9045D

Date/Time 6/12/19 17:10
Received by: J*
Relinquished by: CD

Date/Time 07-12-19 18:10
Received by: CD
Relinquished by: J*