



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus, Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB103954

BALANCE ID: WC SC-4

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP75751
calibration std. phosphate 1 ppm	WP76282
calibration std. phosphate 0.5 ppm	WP76281
calibration std. phosphate 0.3 ppm	WP76280
calibration std. 0 ppm	WP76277
phosphate ICV std.	WP76286
phosphate CCV std.	WP76285
calibration std. phosphate 0.1 ppm	WP76279
calibration std. phosphate 0.05 ppm	WP76278
Combined reagent	WP76284
11 N sulfuric acid	WP71598
Phenolphthalein indicator	WP71599
Sodium hydroxide, 1N	WP71600

Intercept: -0.0009

Slope: 0.6495

Regression: 0.999799

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	07/15/2019	11:03
2	CAL2	0.05	1	50	50	0.033	07/15/2019	11:03
3	CAL3	0.10	1	50	50	0.065	07/15/2019	11:04
4	CAL4	0.30	1	50	50	0.185	07/15/2019	11:05
5	CAL5	0.50	1	50	50	0.330	07/15/2019	11:05
6	CAL6	1.00	1	50	50	0.648	07/15/2019	11:06



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus, Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB103954

BALANCE ID: WC SC-4

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.321	0.496	07/15/2019	11:06
2	ICB		1	50	50	0.000	0.001	07/15/2019	11:07
3	CCV1	0.50	1	50	50	0.324	0.500	07/15/2019	11:08
4	CCB1		1	50	50	0.000	0.001	07/15/2019	11:08
5	PB121378BL		1	1.01	50	0.000	0.001	07/15/2019	11:09
6	PB121378BS	0.50	1	1.02	50	0.329	0.508	07/15/2019	11:09
7	K3591-03		1	1.00	50	0.008	0.014	07/15/2019	11:10
8	CCV2	0.50	1	50	50	0.327	0.505	07/15/2019	11:11
9	CCB2		1	50	50	0.000	0.001	07/15/2019	11:11