



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB103974

Slope : 98.4

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	07/15/2019	17:30
2	CAL2	1	Water	NA	NA	20.1	7.00	07/15/2019	17:31
3	CAL3	1	Water	NA	NA	20.1	10.02	07/15/2019	17:32
4	ICV	1	Water	NA	NA	20.1	6.99	07/15/2019	17:34
5	CCV1	1	Water	NA	NA	20.3	2.02	07/15/2019	18:20
6	K3827-03	1	Water	NA	NA	21.2	6.25	07/15/2019	18:21
7	K3827-03DUP	1	Water	NA	NA	21.1	6.26	07/15/2019	18:22
8	CCV2	1	Water	NA	NA	20.2	12.01	07/15/2019	18:25

LB 103974

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K3827 WorkList ID : 128573 Department : Wet-Chemistry Date : 07-15-2019 16:00:53

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/15/2019	Water	K3827-03-17	pH	Cool 4 deg C	DEWB01	M51	UTILITY-ROOM-BATHROOM	07/15/2019	9040C

Date/Time 07-15-19 16:10
 Received by: JA
 Relinquished by: JA

Date/Time 07-15-19 17:120
 Received by: JA
 Relinquished by: JA