

LB103977

Calibration results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : CT

Instrument ID : Konelab

7/15/2019 14:04

Test Phenolics-


Accepted 7/15/2019 14:03

Factor 9.067

Bias 0.012

Coeff. of det. 0.999664

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.018	0.0523	0.0500	
2	Phenol-2	0.023	0.1052	0.1000	
3	Phenol-2	0.040	0.2530	0.2500	
4	Phenol-2	0.068	0.5049	0.5000	
5	Phenol-2	0.119	0.9729	1.0000	
6	Phenol-2	0.234	2.0116	2.0000	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : CT

Instrument ID : Konelab

7/15/2019 15:57

Test: Phenolics-

Sample Id	Result <i>mg/L</i>	Dil. 1 +	Response	Errors
ICV1	1.009	0.0	0.123	
ICB1	0.002	0.0	0.012	
CCV1	1.012	0.0	0.123	
CCB1	0.000	0.0	0.012	
PB121278BL	0.002	0.0	0.012	
PB121278BS	0.964	0.0	0.118	
K3591-12	0.007	0.0	0.013	
DOC-WATER-01	0.962	0.0	0.118	
DOC-WATER-02	1.013	0.0	0.124	
DOC-WATER-03	0.962	0.0	0.118	
DOC-WATER-04	1.011	0.0	0.123	
CCV2	1.017	0.0	0.124	
CCB2	-0.000	0.0	0.012	

N	13
Mean	0.612
SD	0.5025
CV%	82.07

CT

Aquakem v. 7.2AQ1

Results from time period:

Mon Jul 15 15:41:26 2019

Mon Jul 15 15:44:14 2019

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0523	mg/l	7/15/2019 13:57:00	
0.1PPM	A	Phenolics-	P	0.1052	mg/l	7/15/2019 13:57:01	
0.25PPM	A	Phenolics-	P	0.253	mg/l	7/15/2019 13:57:02	
0.50PPM	A	Phenolics-	P	0.5049	mg/l	7/15/2019 13:57:03	
1.0PPM	A	Phenolics-	P	0.9729	mg/l	7/15/2019 13:57:04	
2.0PPM	A	Phenolics-	P	2.0116	mg/l	7/15/2019 13:57:05	
ICV1	S	Phenolics-	P	1.0091	mg/l	7/15/2019 15:41:26	
ICB1	S	Phenolics-	P	0.0016	mg/l	7/15/2019 15:41:27	
CCV1	S	Phenolics-	P	1.0117	mg/l	7/15/2019 15:41:28	
CCB1	S	Phenolics-	P	0	mg/l	7/15/2019 15:41:29	
PB121278BL	S	Phenolics-	P	0.0023	mg/l	7/15/2019 15:41:30	
PB121278BS	S	Phenolics-	P	0.9638	mg/l	7/15/2019 15:41:31	
K3591-12	S	Phenolics-	P	0.0069	mg/l	7/15/2019 15:41:32	
DOC-WATER-01	S	Phenolics-	P	0.9617	mg/l	7/15/2019 15:41:33	
DOC-WATER-02	S	Phenolics-	P	1.0133	mg/l	7/15/2019 15:41:34	
DOC-WATER-03	S	Phenolics-	P	0.9618	mg/l	7/15/2019 15:41:35	
DOC-WATER-04	S	Phenolics-	P	1.0109	mg/l	7/15/2019 15:41:36	
CCV2	S	Phenolics-	P	1.0165	mg/l	7/15/2019 15:41:37	
CCB2	S	Phenolics-	P	-0.0003	mg/l	7/15/2019 15:44:14	