

Analysis Method: 9034
 Parameter: Reactive Sulfide
 Run Number: LB103982

ANALYST: ketankumar
 SUPERVISOR REVIEW BY: apatel
 Constant: 16000
 Normality1: 0.025
 Normality2: 0.025

Reagent/Standard	Lot/Log #
Sodium thiosulfate titrant 0.025N	WP74519
IODINE SOLUTION .025N 1L	W2480
Starch Solution, 4L	W2379

Seq	Lab ID	True Value (mg/l)	DF	Initial Weight (g)	Final Volume (ml)	T1 (ml)	T2 Initial	T2 Final	T2 Diff. (ml)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	Anal Date	Anal Time
1	PB121387BL		1	5.02	50	2.00	0.00	1.94	1.94	0.06	0.00	0.00	07/15/2019	12:05
2	K3620-02		1	5.03	50	2.00	1.94	3.84	1.90	0.10	0.04	3.18	07/15/2019	12:08
3	K3620-02DUP		1	5.04	50	2.00	3.84	5.72	1.88	0.12	0.06	4.76	07/15/2019	12:12
4	K3620-04		1	5.04	50	2.00	5.72	7.64	1.92	0.08	0.02	1.59	07/15/2019	12:15
5	K3620-06		1	5.05	50	2.00	0.00	1.90	1.90	0.10	0.04	3.17	07/15/2019	12:18
6	K3620-08		1	5.04	50	2.00	1.90	3.66	1.76	0.24	0.18	14.29	07/15/2019	12:22
7	K3620-10		1	5.02	50	2.00	3.66	5.48	1.82	0.18	0.12	9.56	07/15/2019	12:25

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume