



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB103990

Reagent/Standard	Lot/Log #
Residual chlorine std, 0PPM	WP76364
Residual chlorine std, and CCV , 0.4PPM	WP76365
Residual chlorine LOD, 0.05PPM	WP76362
Residual chlorine LOQ, 0.2PPM	WP76363
Residual chlorine ICV-LCS, 0.4PPM	WP76361
Chlorine Calibration std, 0.1ppm	WP76356
Chlorine Calibration std, 0.2ppm	WP76357
Chlorine Calibration std, 0.8ppm	WP76359
Chlorine Calibration std, 0.0ppm	WP76355
Chlorine Calibration std, 1.6ppm	WP76360
Total Chlorine Powder Pillows	W2249

Intercept: 0.0026

Slope: 0.9982

Regression: 0.999978

Seq	Lab ID	True Val (mg/l)	Dilution	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	07/16/2019	10:46
2	CAL2	0.1	1	0.000	0.100	0.100	07/16/2019	10:46
3	CAL3	0.2	1	0.000	0.210	0.210	07/16/2019	10:47
4	CAL4	0.4	1	0.000	0.400	0.400	07/16/2019	10:48
5	CAL5	0.8	1	0.000	0.800	0.800	07/16/2019	10:49
6	CAL6	1.6	1	0.000	1.600	1.600	07/16/2019	10:50



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB103990

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.3900	0.3900	0.3880	07/16/2019	10:52
2	ICB				1	0.0000	0.0000	0.0000	-0.0030	07/16/2019	10:53
3	CCV1			0.4	1	0.0000	0.4000	0.4000	0.3980	07/16/2019	10:53
4	CCB1				1	0.0000	0.0000	0.0000	-0.0030	07/16/2019	10:54
5	LB103990BL	50	50		1	0.0000	0.0000	0.0000	-0.0030	07/16/2019	10:55
6	LB103990BS	50	50	0.4	1	0.0000	0.4000	0.4000	0.3980	07/16/2019	10:56
7	K3591-07	50	50		1	0.0000	0.0500	0.0500	0.0470	07/16/2019	10:57
8	K3591-08	50	50		1	0.0000	0.1100	0.1100	0.1080	07/16/2019	10:59
9	CCV2			0.4	1	0.0000	0.4100	0.4100	0.4080	07/16/2019	10:59
10	CCB2				1	0.0000	0.0000	0.0000	-0.0030	07/16/2019	11:02