

245.1


LB103991

INSTRUMENT ID: CV1

Linear

 μ Abs.:

37371



A= 0.0000e+000

B= 2.6476e-004

C= -1.2149e-002

Rho= 0.9998120

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.015	-0.015	-11	0.000	-11				
0.05	0.050	0.044	-0.006	211	0.0 %	211				
0.2	0.200	0.194	-0.006	779	0.0 %	779				
2.5	2.500	2.479	-0.021	9409	0.0 %	9409				
5.0	5.000	5.020	0.020	19006	0.0 %	19006				
7.5	7.500	7.646	0.146	28924	0.0 %	28924				
10.0	10.000	9.882	-0.118	37371	0.0 %	37371				

LB103991 INSTRUMENT ID :CV1

Method: 245.1

Operator: Admin

Date of Analysis: 16 Jul 2019 11:54:04

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	-11	-	0.0000	245.1	PPB	16 Jul 2019 11:55:27	S	Std
0.05 - 1	50-05	211	-	0.0500	245.1	PPB	16 Jul 2019 11:57:45	S	Std
0.2 - 1	50-2	779	-	0.2000	245.1	PPB	16 Jul 2019 12:00:03	S	Std
2.5 - 1	52-5	9409	-	2.5000	245.1	PPB	16 Jul 2019 12:02:21	S	Std
5.0 - 1	55-0	19006	-	5.0000	245.1	PPB	16 Jul 2019 12:04:38	S	Std
7.5 - 1	57-5	28924	-	7.5000	245.1	PPB	16 Jul 2019 12:06:55	S	Std
10.0 - 1	510-0	37371	-	10.0000	245.1	PPB	16 Jul 2019 12:16:34	S	Std
ICV33 - 1	ICV33	15465	4.0824	-245.1		PPB	16 Jul 2019 12:39:24	U	SMPL
ICB33 - 1	ICB33	-61	-0.0283	-245.1		PPB	16 Jul 2019 12:41:39	U	SMPL
CCV27 - 1	CCV27	19150	5.0581	-245.1		PPB	16 Jul 2019 12:43:56	U	SMPL
CCB27 - 1	CCB27	-74	-0.0317	-245.1		PPB	16 Jul 2019 12:46:11	U	SMPL
CRA - 1	CRA	217	0.0453	-245.1		PPB	16 Jul 2019 12:48:28	U	SMPL
HighStd - 1	HighStd	37330	9.8715	-245.1		PPB	16 Jul 2019 12:50:44	U	SMPL
ChkStd - 1	ChkStd	26456	6.9925	-245.1		PPB	16 Jul 2019 12:53:00	U	SMPL
PB121430BL - 1	PBW	-46	-0.0243	-245.1		PPB	16 Jul 2019 12:55:18	U	SMPL
PB121430BS - 1	LCSW	15251	4.0258	-245.1		PPB	16 Jul 2019 12:57:35	U	SMPL
K3672-01 - 1	COMP	345	0.0792	-245.1		PPB	16 Jul 2019 12:59:52	U	SMPL
K3672-01DUP - 1	COMPDUP	369	0.0855	-245.1		PPB	16 Jul 2019 13:02:10	U	SMPL
K3672-01MS - 1	COMPMS	15583	4.1137	-245.1		PPB	16 Jul 2019 13:04:27	U	SMPL
K3672-01MSD - 1	COMPMSD	15663	4.1349	-245.1		PPB	16 Jul 2019 13:06:45	U	SMPL
K3672-01LX5 - 1		20	-0.0069	-245.1		PPB	16 Jul 2019 13:09:01	U	SMPL
CCV28 - 1	CCV28	19048	5.0311	-245.1		PPB	16 Jul 2019 13:11:17	U	SMPL
CCB28 - 1	CCB28	-45	-0.0241	-245.1		PPB	16 Jul 2019 13:13:33	U	SMPL
K3672-01A - 1		15755	4.1592	-245.1		PPB	16 Jul 2019 13:15:49	U	SMPL
CCV29 - 1	CCV29	19258	5.0867	-245.1		PPB	16 Jul 2019 13:18:05	U	SMPL
CCB29 - 1	CCB29	14	-0.0084	-245.1		PPB	16 Jul 2019 13:20:21	U	SMPL