



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB103997

Slope : 98.7

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	07/16/2019	13:00
2	CAL2	1	Water	NA	NA	20.1	7.00	07/16/2019	13:01
3	CAL3	1	Water	NA	NA	20.1	10.03	07/16/2019	13:02
4	ICV	1	Water	NA	NA	20.1	6.99	07/16/2019	13:05
5	CCV1	1	Water	NA	NA	20.1	2.02	07/16/2019	13:06
6	K3839-01	1	Water	NA	NA	21.4	5.61	07/16/2019	13:07
7	K3839-01DUP	1	Water	NA	NA	21.3	5.62	07/16/2019	13:08
8	K3839-02	1	Water	NA	NA	21.3	5.09	07/16/2019	13:10
9	CCV2	1	Water	NA	NA	20.2	12.02	07/16/2019	13:11

LD 103997

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K3839 WorkList ID : 128613 Department : Wet-Chemistry Date : 07-16-2019 12:45:36

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/11/2019	Water	K3839-01 - 1	pH	Cool 4 deg C	STER16	M33	MAIN-CELL	07/11/2019	SM 4500-H B
07/11/2019	Water	K3839-02 - 2	pH	Cool 4 deg C	STER16	M33	EX-CELL	07/11/2019	SM 4500-H B

Date/Time 07-16-19 12:50
Received by: JF
Relinquished by: JF

Date/Time 07-16-19 14:00
Received by: PC
Relinquished by: JF