



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB104003

Slope : 98.8

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	07/16/2019	16:00
2	CAL2	1	Water	NA	NA	21.2	7.00	07/16/2019	16:01
3	CAL3	1	Water	NA	NA	21.2	10.03	07/16/2019	16:02
4	ICV	1	Water	NA	NA	21.2	6.99	07/16/2019	16:04
5	CCV1	1	Water	NA	NA	21.2	2.02	07/16/2019	16:05
6	K3838-04	1	Water	NA	NA	26.5	8.24	07/16/2019	16:06
7	K3838-04DUP	1	Water	NA	NA	26.6	8.25	07/16/2019	16:07
8	CCV2	1	Water	NA	NA	21.2	12.02	07/16/2019	16:10

LB 104003

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity k3838 WorkList ID : 128621 Department : Wet-Chemistry Date : 07-16-2019 15:43:43

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/16/2019	Water	K3838-04-F	Corrosivity	Cool 4 deg C	PSEG01	M51	WARINANCO-CITY-SEWER-I	07/16/2019	9040C

Date/Time 7-16-19 15:50
Received by: [Signature]
Relinquished by: [Signature]

Date/Time 07-16-19 17:00
Received by: [Signature]
Relinquished by: [Signature]